



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB100698

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.1	4.01	02/06/2019	17:20
2	CAL2	1	Water	NA	NA	22.1	7.01	02/06/2019	17:21
3	CAL3	1	Water	NA	NA	22.1	10.03	02/06/2019	17:22
4	ICV	1	Water	NA	NA	22.1	7.00	02/06/2019	17:24
5	CCV1	1	Water	NA	NA	22.1	2.02	02/06/2019	17:45
6	K1385-08	1	Solid	20.03	20	21.5	10.37	02/06/2019	17:46
7	K1385-08DUP	1	Solid	20.04	20	21.6	10.38	02/06/2019	17:48
8	K1389-04	1	Solid	20.02	20	21.3	10.68	02/06/2019	17:49
9	K1389-08	1	Solid	20.03	20	21.4	10.51	02/06/2019	17:50
10	K1389-12	1	Solid	20.04	20	21.2	9.98	02/06/2019	17:51
11	K1389-16	1	Solid	20.03	20	20.9	8.85	02/06/2019	18:00
12	K1389-20	1	Solid	20.02	20	20.8	7.30	02/06/2019	18:01
13	K1394-01	1	Solid	20.03	20	21.5	10.83	02/06/2019	18:02
14	K1394-03	1	Solid	20.04	20	21.8	10.51	02/06/2019	18:03
15	K1401-02	1	Solid	20.02	20	21.3	7.09	02/06/2019	18:05
16	CCV2	1	Water	NA	NA	22.1	12.01	02/06/2019	18:06
17	K1401-04	1	Solid	20.04	20	21.5	5.58	02/06/2019	18:09
18	K1401-06	1	Solid	20.01	20	20.9	6.80	02/06/2019	18:10
19	K1401-08	1	Solid	20.03	20	21.0	6.85	02/06/2019	18:11
20	K1401-10	1	Solid	20.03	20	21.1	5.56	02/06/2019	18:13

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	CCV3	1	Water	NA	NA	21.1	2.01	02/06/2019	18:15

VB 100698

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity K1385

WorkList ID : 122019

Date : 02-06-2019 15:08:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/05/2019	Solid	K1385-08	Corrosivity	Cool 4 deg C	PARS15	L43	98-WC-S-01-2-57	02/05/2019	9045D
02/06/2019	Solid	K1389-04	Corrosivity	Cool 4 deg C	PSEG01	K31	M-4	02/06/2019	9045D
02/06/2019	Solid	K1389-08	Corrosivity	Cool 4 deg C	PSEG01	K31	M-6	02/06/2019	9045D
02/06/2019	Solid	K1389-12	Corrosivity	Cool 4 deg C	PSEG01	K31	M-5S	02/06/2019	9045D
02/06/2019	Solid	K1389-16	Corrosivity	Cool 4 deg C	PSEG01	K31	M-5-D	02/06/2019	9045D
02/06/2019	Solid	K1389-20	Corrosivity	Cool 4 deg C	PSEG01	K31	M-7-S	02/06/2019	9045D
02/06/2019	Solid	K1394-01	Corrosivity	Cool 4 deg C	PSEG01	K31	ISW	02/06/2019	9045D
02/06/2019	Solid	K1394-03	Corrosivity	Cool 4 deg C	PSEG01	K31	ISF	02/06/2019	9045D
02/06/2019	Solid	K1401-02	Corrosivity	Cool 4 deg C	PSEG01	K31	RBR-200050	02/06/2019	9045D
02/06/2019	Solid	K1401-04	Corrosivity	Cool 4 deg C	PSEG01	K31	RBR-200028	02/06/2019	9045D
02/06/2019	Solid	K1401-06	Corrosivity	Cool 4 deg C	PSEG01	K31	RT-3425	02/06/2019	9045D
02/06/2019	Solid	K1401-08	Corrosivity	Cool 4 deg C	PSEG01	K31	RT-3451	02/06/2019	9045D
02/06/2019	Solid	K1401-10	Corrosivity	Cool 4 deg C	PSEG01	K31	RBR-200052	02/06/2019	9045D

Date/Time 02-06-19 15:25:27
 Received by: JS
 Relinquished by: JS

Date/Time 02-06-19 7:00:00
 Received by: JS
 Relinquished by: JS