


LB100708
INSTRUMENT ID: CV1

7471B

Linear



A= 0.0000e+000

B= 1.7664e-004

C= -4.3452e-002

Rho= 0.9999530

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.041	-0.041	14	0.000	14				
0.2	0.200	0.182	-0.018	1279	0.0 %	1279				
2.5	2.500	2.562	0.062	14749	0.0 %	14749				
5.0	5.000	5.017	0.017	28648	0.0 %	28648				
7.5	7.500	7.516	0.016	42794	0.0 %	42794				
10.0	10.000	9.964	-0.036	56651	0.0 %	56651				

LB100708 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 07 Feb 2019 10:00:26

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	14	-	0.0000	7471B	PPB	07 Feb 2019 10:21:47	S	Std
0.2 - 1	50-2	1279	-	0.2000	7471B	PPB	07 Feb 2019 10:24:04	S	Std
2.5 - 1	52-5	14749	-	2.5000	7471B	PPB	07 Feb 2019 10:26:21	S	Std
5.0 - 1	55-0	28648	-	5.0000	7471B	PPB	07 Feb 2019 10:28:37	S	Std
7.5 - 1	57-5	42794	-	7.5000	7471B	PPB	07 Feb 2019 10:30:55	S	Std
10.0 - 1	510-0	56651	-	10.0000	7471B	PPB	07 Feb 2019 10:33:12	S	Std
ICV99 - 1	ICV99	22202	3.8784	-	7471B	PPB	07 Feb 2019 10:36:53	U	SMPL
ICB99 - 1	ICB99	-80	-0.0576	-	7471B	PPB	07 Feb 2019 10:39:10	U	SMPL
CCV94 - 1	CCV94	28664	5.0199	-	7471B	PPB	07 Feb 2019 10:41:26	U	SMPL
CCB94 - 1	CCB94	-97	-0.0606	-	7471B	PPB	07 Feb 2019 10:43:41	U	SMPL
CRA - 1	CRA	1405	0.2047	-	7471B	PPB	07 Feb 2019 10:50:08	U	SMPL
HighStd - 1	HighStd	57005	10.0262	-	7471B	PPB	07 Feb 2019 10:52:23	U	SMPL
ChkStd - 1	ChkStd	39039	6.8526	-	7471B	PPB	07 Feb 2019 10:54:39	U	SMPL
PB116906BL - 1	PBS	143	-0.0182	-	7471B	PPB	07 Feb 2019 10:56:56	U	SMPL
PB116906BS - 1	LCSS	19886	3.4693	-	7471B	PPB	07 Feb 2019 10:59:14	U	SMPL
K1383-01 - 1	A0C5B-SS1-0.0-0.5	3321	0.5432	-	7471B	PPB	07 Feb 2019 11:01:31	U	SMPL
K1383-02 - 1	A0C5B-SS1-4.5-5.0	1747	0.2651	-	7471B	PPB	07 Feb 2019 11:03:48	U	SMPL
K1383-03 - 1	A0C5B-SS2-0.0-0.5	2785	0.4485	-	7471B	PPB	07 Feb 2019 11:06:05	U	SMPL
K1385-01 - 1	98-SB-01-2-4	8597	1.4752	-	7471B	PPB	07 Feb 2019 11:08:22	U	SMPL
K1385-01DUP - 1	98-SB-01-2-4DUP	8540	1.4651	-	7471B	PPB	07 Feb 2019 11:10:37	U	SMPL
CCV95 - 1	CCV95	28571	5.0035	-	7471B	PPB	07 Feb 2019 11:12:53	U	SMPL
CCB95 - 1	CCB95	-74	-0.0565	-	7471B	PPB	07 Feb 2019 11:15:08	U	SMPL
K1385-02 - 1	K1385-01MS	28497	4.9904	-	7471B	PPB	07 Feb 2019 11:17:25	U	SMPL
K1385-03 - 1	K1385-01MSD	28859	5.0543	-	7471B	PPB	07 Feb 2019 11:19:40	U	SMPL
K1385-04 - 1	98-SB-101-2-4	7170	1.2231	-	7471B	PPB	07 Feb 2019 11:21:57	U	SMPL
K1385-06 - 1	98-SB-01-4-57	535	0.0511	-	7471B	PPB	07 Feb 2019 11:24:14	U	SMPL
K1389-03 - 1	M-4	873	0.1108	-	7471B	PPB	07 Feb 2019 11:26:30	U	SMPL
K1389-07 - 1	M-6	3250	0.5306	-	7471B	PPB	07 Feb 2019 11:28:47	U	SMPL
K1389-11 - 1	M-5S	1140	0.1579	-	7471B	PPB	07 Feb 2019 11:31:05	U	SMPL
K1389-15 - 1	M-5-D	436	0.0336	-	7471B	PPB	07 Feb 2019 11:33:22	U	SMPL
K1389-19 - 1	M-7-S	518	0.0480	-	7471B	PPB	07 Feb 2019 11:35:39	U	SMPL
K1401-01 - 1	RBR-200050	1815	0.2772	-	7471B	PPB	07 Feb 2019 11:37:55	U	SMPL
CCV96 - 1	CCV96	28587	5.0063	-	7471B	PPB	07 Feb 2019 11:40:11	U	SMPL
CCB96 - 1	CCB96	-79	-0.0574	-	7471B	PPB	07 Feb 2019 11:42:26	U	SMPL
K1401-03 - 1	RBR-200028	1963	0.3033	-	7471B	PPB	07 Feb 2019 11:44:43	U	SMPL
K1401-05 - 1	RT-3425	3088	0.5020	-	7471B	PPB	07 Feb 2019 11:46:58	U	SMPL
K1401-07 - 1	RT-3451	7038	1.1998	-	7471B	PPB	07 Feb 2019 11:49:14	U	SMPL
K1401-09 - 1	RBR-200052	1034	0.1392	-	7471B	PPB	07 Feb 2019 11:51:30	U	SMPL
K1403-01 - 1	S-1	2395	0.3796	-	7471B	PPB	07 Feb 2019 11:53:46	U	SMPL
K1385-01LX5 - 1		1624	0.2434	-	7471B	PPB	07 Feb 2019 11:56:02	U	SMPL
K1385-01A - 1		30674	5.3750	-	7471B	PPB	07 Feb 2019 11:58:18	U	SMPL
CCV97 - 1	CCV97	27912	4.8871	-	7471B	PPB	07 Feb 2019 12:00:35	U	SMPL
CCB97 - 1	CCB97	-78	-0.0572	-	7471B	PPB	07 Feb 2019 12:02:53	U	SMPL