


LB100717

INSTRUMENT ID : CV1

7470A

μ Abs.:

41342



Linear

A= 0.0000e+000

B= 2.4064e-004

C= -1.5353e-002

Rho= 0.9999163

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.040	-0.040	-102	0.000	-102				
0.2	0.200	0.178	-0.022	802	0.0 %	802				
2.5	2.500	2.524	0.024	10553	0.0 %	10553				
5.0	5.000	5.075	0.075	21155	0.0 %	21155				
7.5	7.500	7.529	0.029	31352	0.0 %	31352				
10.0	10.000	9.933	-0.067	41342	0.0 %	41342				

LB100717 INSTRUMENT ID :CV1

Method: 7470A

Operator: Admin

Date of Analysis: 08 Feb 2019 11:13:17

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	-102	-	0.0000	7470A	PPB	08 Feb 2019 12:18:59	S	Std
0.2 - 1	50-2	802	-	0.2000	7470A	PPB	08 Feb 2019 12:21:08	S	Std
2.5 - 1	52-5	10553	-	2.5000	7470A	PPB	08 Feb 2019 12:23:15	S	Std
5.0 - 1	55-0	21155	-	5.0000	7470A	PPB	08 Feb 2019 12:25:23	S	Std
7.5 - 1	57-5	31352	-	7.5000	7470A	PPB	08 Feb 2019 12:27:30	S	Std
10.0 - 1	510-0	41342	-	10.0000	7470A	PPB	08 Feb 2019 12:29:37	S	Std
ICV100 - 1	ICV100	17734	4.2522	-	7470A	PPB	08 Feb 2019 12:32:32	U	SMPL
ICB100 - 1	ICB100	-52	-0.0279	-	7470A	PPB	08 Feb 2019 12:34:39	U	SMPL
CCV98 - 1	CCV98	21136	5.0709	-	7470A	PPB	08 Feb 2019 12:36:45	U	SMPL
CCB98 - 1	CCB98	15	-0.0117	-	7470A	PPB	08 Feb 2019 12:43:46	U	SMPL
CRA - 1	CRA	827	0.1837	-	7470A	PPB	08 Feb 2019 12:45:51	U	SMPL
HighStd - 1	HighStd	41128	9.8818	-	7470A	PPB	08 Feb 2019 12:51:10	U	SMPL
ChkStd - 1	ChkStd	29963	7.1950	-	7470A	PPB	08 Feb 2019 12:57:14	U	SMPL
PB116883BL - 1	PBW	115	0.0123	-	7470A	PPB	08 Feb 2019 12:59:20	U	SMPL
PB116883BS - 1	LCSW	16001	3.8352	-	7470A	PPB	08 Feb 2019 13:01:28	U	SMPL
K1338-02 - 1	AU-06-020519-A	177	0.0272	-	7470A	PPB	08 Feb 2019 13:03:35	U	SMPL
K1338-02DUP - 1	AU-06-020519-ADUP	210	0.0352	-	7470A	PPB	08 Feb 2019 13:10:06	U	SMPL
K1338-02MS - 1	AU-06-020519-AMS	16053	3.8477	-	7470A	PPB	08 Feb 2019 13:12:13	U	SMPL
K1338-02MSD - 1	AU-06-020519-AMSD	16128	3.8657	-	7470A	PPB	08 Feb 2019 13:14:21	U	SMPL
K1338-04 - 1	AU-06-020519-B	134	0.0169	-	7470A	PPB	08 Feb 2019 13:16:27	U	SMPL
CCV99 - 1	CCV99	21147	5.0735	-	7470A	PPB	08 Feb 2019 13:18:33	U	SMPL
CCB99 - 1	CCB99	-25	-0.0214	-	7470A	PPB	08 Feb 2019 13:20:40	U	SMPL
K1338-06 - 1	AU-06-020519-C	176	0.0270	-	7470A	PPB	08 Feb 2019 13:22:46	U	SMPL
K1356-02 - 1	SU-01-020519-A	189	0.0301	-	7470A	PPB	08 Feb 2019 13:24:52	U	SMPL
K1356-04 - 1	SU-01-020519-B	210	0.0352	-	7470A	PPB	08 Feb 2019 13:26:58	U	SMPL
K1358-02 - 1	TR-05-020519	261	0.0475	-	7470A	PPB	08 Feb 2019 13:29:05	U	SMPL
K1379-02 - 1	HR-6A	450	0.0929	-	7470A	PPB	08 Feb 2019 13:31:12	U	SMPL
PB116907BL - 1	PBW	137	0.0176	-	7470A	PPB	08 Feb 2019 13:36:38	U	SMPL
PB116907BS - 1	LCSW	15583	3.7346	-	7470A	PPB	08 Feb 2019 13:38:45	U	SMPL
K1383-04 - 1	FB-02-05-19	211	0.0354	-	7470A	PPB	08 Feb 2019 13:40:53	U	SMPL
K1389-21 - 1	M-7	217	0.0369	-	7470A	PPB	08 Feb 2019 13:43:00	U	SMPL
K1392-01 - 1	MW-11	310	0.0592	-	7470A	PPB	08 Feb 2019 13:45:07	U	SMPL
CCV01 - 1	CCV01	21091	5.0601	-	7470A	PPB	08 Feb 2019 13:47:14	U	SMPL
CCB01 - 1	CCB01	-60	-0.0298	-	7470A	PPB	08 Feb 2019 13:49:20	U	SMPL
K1392-01DUP - 1	MW-11DUP	308	0.0588	-	7470A	PPB	08 Feb 2019 13:51:26	U	SMPL
K1392-01MS - 1	MW-11MS	15107	3.6200	-	7470A	PPB	08 Feb 2019 13:53:32	U	SMPL
K1392-01MSD - 1	MW-11MSD	14843	3.5565	-	7470A	PPB	08 Feb 2019 13:55:39	U	SMPL
K1392-03 - 1	MW-9	481	0.1004	-	7470A	PPB	08 Feb 2019 13:57:45	U	SMPL
K1392-04 - 1	MW-8	370	0.0737	-	7470A	PPB	08 Feb 2019 13:59:52	U	SMPL
K1392-05 - 1	MW-7	507	0.1067	-	7470A	PPB	08 Feb 2019 14:01:59	U	SMPL
K1392-06 - 1	MW-6	454	0.0939	-	7470A	PPB	08 Feb 2019 14:04:07	U	SMPL
K1392-07 - 1	FIELD-BLANK	171	0.0258	-	7470A	PPB	08 Feb 2019 14:06:15	U	SMPL
K1392-08 - 1	DUP-1	354	0.0698	-	7470A	PPB	08 Feb 2019 14:08:22	U	SMPL
PB116883TB - 1	PB116859TB	202	0.0333	-	7470A	PPB	08 Feb 2019 14:10:29	U	SMPL
CCV02 - 1	CCV02	21129	5.0692	-	7470A	PPB	08 Feb 2019 14:12:36	U	SMPL
CCB02 - 1	CCB02	-56	-0.0288	-	7470A	PPB	08 Feb 2019 14:14:42	U	SMPL
K1338-02LX5 - 1		26	-0.0091	-	7470A	PPB	08 Feb 2019 14:16:49	U	SMPL
K1338-02A - 1		16434	3.9394	-	7470A	PPB	08 Feb 2019 14:18:55	U	SMPL
K1392-01LX5 - 1		44	-0.0048	-	7470A	PPB	08 Feb 2019 14:21:02	U	SMPL
K1392-01A - 1		16611	3.9820	-	7470A	PPB	08 Feb 2019 14:23:09	U	SMPL
CCV03 - 1	CCV03	21120	5.0670	-	7470A	PPB	08 Feb 2019 14:25:15	U	SMPL
CCB03 - 1	CCB03	-54	-0.0283	-	7470A	PPB	08 Feb 2019 14:27:23	U	SMPL