



Analytical Summary Report

Analysis Method: 9040C
Parameter: Corrosivity
Run Number: LB100724

Analyst By : JIGNESH
Supervisor Review By : apatel
Slope : 98.3
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	02/07/2019	12:20
2	CAL2	1	Water	NA	NA	21.1	7.01	02/07/2019	12:21
3	CAL3	1	Water	NA	NA	21.1	10.02	02/07/2019	12:22
4	ICV	1	Water	NA	NA	21.1	6.99	02/07/2019	12:24
5	CCV1	1	Water	NA	NA	21.2	2.02	02/07/2019	12:35
6	K1405-01	1	Water	NA	NA	21.5	4.13	02/07/2019	12:36
7	K1406-01	1	Water	NA	NA	20.8	7.61	02/07/2019	12:37
8	K1406-01DUP	1	Water	NA	NA	20.7	7.62	02/07/2019	12:39
9	CCV2	1	Water	NA	NA	21.1	12.01	02/07/2019	12:40

LB100724

WORKLIST(Hardcopy Internal Chain)

02-07-19 12:20

WorkList Name : CORROSIVITY W K1405 WorkList ID : 122116 Date : 02-08-2019 12:19:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/06/2019	Water	K1405-01	Corrosivity	Cool 4 deg C	WARR01	L51	26-WARD-UNK-LIQUID	02/06/2019	9040C
02/06/2019	Water	K1406-01	Corrosivity	Cool 4 deg C	WARR01	L51	N.RIVER-UNK-LIQUID	02/06/2019	9040C

Date/Time 02.07.19 12.25
 Received by: Jr
 Relinquished by: CP

Date/Time 02-07-19 10:00 AM
 Received by: CP
 Relinquished by: Jr