



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB100725

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	02/07/2019	12:20
2	CAL2	1	Water	NA	NA	21.1	7.01	02/07/2019	12:21
3	CAL3	1	Water	NA	NA	21.1	10.02	02/07/2019	12:22
4	ICV	1	Water	NA	NA	21.1	6.99	02/07/2019	12:24
5	CCV1	1	Water	NA	NA	21.2	2.02	02/07/2019	12:25
6	K1404-01	1	Water	NA	NA	17.8	6.45	02/07/2019	12:26
7	K1414-01	1	Water	NA	NA	16.9	5.23	02/07/2019	12:27
8	K1414-01DUP	1	Water	NA	NA	17.0	5.24	02/07/2019	12:30
9	CCV2	1	Water	NA	NA	21.2	12.02	02/07/2019	12:31

203 100725

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K1414

WorkList ID : 122113

02-07-19

Date : 02-08-2019 12:10:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/06/2019	Water	K1404-01-YA	pH	Cool 4 deg C	WEST10	L61	FB-02-06-19	02/06/2019	SM 4500-H B
02/05/2019	Water	K1414-01-f)	pH	Cool 4 deg C	STER06	L33	POOL-WATER	02/05/2019	SM 4500-H B

02-07-19

02-08-19 12:15

Date/Time

Received by:

Relinquished by:

30

Date/Time

Received by:

Relinquished by:

02-07-19

21:00 PM

60

30