



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB100738

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.9	02/07/2019	10:10
2	K1389-22		1	8	>100.0	>212.0	>212.0	02/07/2019	10:40
3	K1389-22DUP		1	8	>100.0	>212.0	>212.0	02/07/2019	11:10
4	K1396-02		1	9	>100.0	>212.0	>212.0	02/07/2019	11:40
5	K1396-03		1	7	>100.0	>212.0	>212.0	02/07/2019	12:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB100738

WorkList Name : FP2719

WorkList ID : 122058

Date : 02-07-2019 09:24:45

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/11/2019	Water	K1389-22	F Flash Point	Cool 4 deg C	PSEG01	K31	M-7	02/06/2019	1010A
02/11/2019	Water	K1396-02	C Flash Point	Cool 4 deg C	PSEG01	K31	COMP-1	02/06/2019	1010A
02/11/2019	Water	K1396-03	A Flash Point	Cool 4 deg C	PSEG01	K31	COMP-2	02/06/2019	1010A

Date/Time 02/07/19 9:25
 Received by: KS
 Relinquished by: CP

Date/Time 02/07/19 12:20
 Received by: CP
 Relinquished by: KS