

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB100767

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP70044
calibration std. phosphate 1 ppm	WP71535
calibration std. phosphate 0.5 ppm	WP71534
calibration std. phosphate 0.3 ppm	WP71533
calibration std. 0 ppm	WP71530
phosphate ICV std.	WP71539
phosphate CCV std.	WP71538
calibration std. phosphate 0.1 ppm	WP71532
calibration std. phosphate 0.05 ppm	WP71531
Combined reagent	WP71597
11 N sulfuric acid	WP71598
Phenolphthalein indicator	WP71599
Sodium hydroxide, 1N	WP71600

Intercept: -0.001Slope: 0.6502Regression: 0.999847

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	02/11/2019	11:30
2	CAL2	0.05	1	50	50	0.033	02/11/2019	11:30
3	CAL3	0.10	1	50	50	0.065	02/11/2019	11:31
4	CAL4	0.30	1	50	50	0.186	02/11/2019	11:31
5	CAL5	0.50	1	50	50	0.329	02/11/2019	11:32
6	CAL6	1.00	1	50	50	0.649	02/11/2019	11:33



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB100767

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.325	0.501	02/11/2019	11:34
2	ICB		1	50	50	0.000	0.002	02/11/2019	11:34
3	CCV1	0.50	1	50	50	0.327	0.504	02/11/2019	11:35
4	CCB1		1	50	50	0.000	0.002	02/11/2019	11:36
5	PB117011BL		1	50	50	0.000	0.002	02/11/2019	11:36
6	PB117011BS	0.50	1	50	50	0.321	0.495	02/11/2019	11:37
7	K1421-06		1	50	50	0.177	0.274	02/11/2019	11:37
8	K1434-01		1	50	50	1.424	2.192	02/11/2019	11:38
9	K1437-01		1	50	50	0.010	0.017	02/11/2019	11:38
10	K1437-01DUP		1	50	50	0.009	0.015	02/11/2019	11:39
11	K1437-01MS	0.50	1	50	50	0.317	0.489	02/11/2019	11:40
12	K1437-01MSD	0.50	1	50	50	0.320	0.494	02/11/2019	11:40
13	K1434-01		10	50	50	0.377	0.581	02/11/2019	11:46
14	CCV2	0.50	1	50	50	0.326	0.503	02/11/2019	11:47
15	CCB2		1	50	50	0.000	0.002	02/11/2019	11:47