

20100774

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

2/11/2019 14:26

Test: Total CN

Sample Id	<i>Mg/d</i> Result	Dil. 1 +	Response	Errors
ICV1	99.289	0.0	0.097	
ICB1	-0.486	0.0	0.004	
CCV1	251.145	0.0	0.238	
CCB1	-0.538	0.0	0.004	
PB116979BL	-0.391	0.0	0.004	
PB116979BS	100.494	0.0	0.098	
MIDPB116979	253.330	0.0	0.240	
K1443-01MS	26.278	0.0	0.029	
K1443-01MSD	26.055	0.0	0.029	
K1443-01	0.742	0.0	0.005	
K1443-01DUP	0.185	0.0	0.005	
CCV2	267.626	0.0	0.254	
CCB2	-0.599	0.0	0.004	

N	13
Mean	78.702
SD	107.8886
CV%	137.08

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : 

2/11/2019 14:37


Test Total CN

Accepted 2/11/2019 9:54

Factor 1073
Bias 0.004

Coeff. of det. 0.999919

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.8765	0.0000	
2	5.0PPBCN	0.008	4.1553	5.0000	
3	10PPBCN	0.013	9.4480	10.0000	
4	50PPBCN	0.051	49.9046	50.0000	
5	100PPBCN	0.098	100.8612	100.0000	
6	250PPBCN	0.240	253.3013	250.0000	
7	500PPBCN	0.469	498.2061	500.0000	

Ag

Aquakem v. 7.2AQ1

Results from time period:

Mon Feb 11 14:00:57 2019

Mon Feb 11 14:21:47 2019

Sample Id	Sam/Ctr/c	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-0.8765	µg/l	2/11/2019 9:53:08	
5.0PPBCN	A	Total CN	P	4.1553	µg/l	2/11/2019 9:53:09	
10PPBCN	A	Total CN	P	9.448	µg/l	2/11/2019 9:53:10	
50PPBCN	A	Total CN	P	49.9046	µg/l	2/11/2019 9:53:11	
100PPBCN	A	Total CN	P	100.8612	µg/l	2/11/2019 9:53:12	
250PPBCN	A	Total CN	P	253.3013	µg/l	2/11/2019 9:53:13	
500PPBCN	A	Total CN	P	498.2061	µg/l	2/11/2019 9:53:14	
ICV1	S	Total CN	P	99.2887	µg/l	2/11/2019 14:00:57	
ICB1	S	Total CN	P	-0.4865	µg/l	2/11/2019 14:00:58	
CCV1	S	Total CN	P	251.1454	µg/l	2/11/2019 14:00:59	
CCB1	S	Total CN	P	-0.538	µg/l	2/11/2019 14:01:00	
PB116979BL	S	Total CN	P	-0.3909	µg/l	2/11/2019 14:01:01	
PB116979BS	S	Total CN	P	100.4939	µg/l	2/11/2019 14:01:02	
MIDPB116979	S	Total CN	P	253.3297	µg/l	2/11/2019 14:01:03	
K1443-01MS	S	Total CN	P	26.2783	µg/l	2/11/2019 14:01:06	
K1443-01MSD	S	Total CN	P	26.0549	µg/l	2/11/2019 14:01:07	
K1443-01	S	Total CN	P	0.7416	µg/l	2/11/2019 14:21:44	
K1443-01DUP	S	Total CN	P	0.1849	µg/l	2/11/2019 14:21:45	
CCV2	S	Total CN	P	267.6259	µg/l	2/11/2019 14:21:46	
CCB2	S	Total CN	P	-0.599	µg/l	2/11/2019 14:21:47	