

LB100775


INSTRUMENT ID: CV1

7471B

Linear

μ Abs.:

56788



A= 0.0000e+000

B= 1.7453e-004

C= 1.3907e-003

Rho= 0.9992515

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.029	0.029	157	0.000	157				
0.2	0.200	0.212	0.012	1209	0.0 %	1209				
2.5	2.500	2.535	0.035	14517	0.0 %	14517				
5.0	5.000	4.769	-0.231	27316	0.0 %	27316				
7.5	7.500	7.743	0.243	44356	0.0 %	44356				
10.0	10.000	9.912	-0.088	56788	0.0 %	56788				

LB100775 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 11 Feb 2019 14:51:07

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	157	-	0.0000	7471B	PPB	11 Feb 2019 14:52:31	S	Std
0.2 - 1	50.2	1209	-	0.2000	7471B	PPB	11 Feb 2019 14:54:48	S	Std
2.5 - 1	52.5	14517	-	2.5000	7471B	PPB	11 Feb 2019 14:57:05	S	Std
5.0 - 1	55.0	27316	-	5.0000	7471B	PPB	11 Feb 2019 14:59:22	S	Std
7.5 - 1	57.5	44356	-	7.5000	7471B	PPB	11 Feb 2019 15:04:50	S	Std
10.0 - 1	510.0	56788	-	10.0000	7471B	PPB	11 Feb 2019 15:07:05	S	Std
ICV05 - 1	ICV05	22698	3.9628	-	7471B	PPB	11 Feb 2019 15:11:30	U	SMPL
ICB05 - 1	ICB05	26	0.0059	-	7471B	PPB	11 Feb 2019 15:13:45	U	SMPL
CCV17 - 1	CCV17	27810	4.8550	-	7471B	PPB	11 Feb 2019 15:16:00	U	SMPL
CCB17 - 1	CCB17	35	0.0075	-	7471B	PPB	11 Feb 2019 15:18:15	U	SMPL
CRA - 1	CRA	1167	0.2051	-	7471B	PPB	11 Feb 2019 15:20:32	U	SMPL
HighStd - 1	HighStd	57177	9.9803	-	7471B	PPB	11 Feb 2019 15:22:47	U	SMPL
ChkStd - 1	ChkStd	40418	7.0554	-	7471B	PPB	11 Feb 2019 15:25:03	U	SMPL
PB117008BL - 1	PBS	-100	-0.0161	-	7471B	PPB	11 Feb 2019 15:27:21	U	SMPL
PB117008BS - 1	LCSS	20619	3.5999	-	7471B	PPB	11 Feb 2019 15:29:39	U	SMPL
K1348-01 - 1	OK-01-020719	1894	0.3319	-	7471B	PPB	11 Feb 2019 15:31:55	U	SMPL
K1348-01DUP - 1	OK-01-020719DUP	1920	0.3365	-	7471B	PPB	11 Feb 2019 15:34:12	U	SMPL
K1348-01MS - 1	OK-01-020719MS	23273	4.0631	-	7471B	PPB	11 Feb 2019 15:36:29	U	SMPL
K1348-01MSD - 1	OK-01-020719MSD	23050	4.0242	-	7471B	PPB	11 Feb 2019 15:38:46	U	SMPL
K1349-01 - 1	OR-03-020719-A	1923	0.3370	-	7471B	PPB	11 Feb 2019 15:41:01	U	SMPL
CCV18 - 1	CCV18	27558	4.8110	-	7471B	PPB	11 Feb 2019 15:43:17	U	SMPL
CCB18 - 1	CCB18	41	0.0085	-	7471B	PPB	11 Feb 2019 15:45:32	U	SMPL
K1349-03 - 1	OR-03-020719-B	2144	0.3756	-	7471B	PPB	11 Feb 2019 15:47:49	U	SMPL
K1356-05 - 1	SU-03-020819-A	2224	0.3895	-	7471B	PPB	11 Feb 2019 15:50:04	U	SMPL
K1356-07 - 1	SU-03-020819-B	1823	0.3196	-	7471B	PPB	11 Feb 2019 15:52:20	U	SMPL
K1422-01 - 1	TEST-PIT-1	1057	0.1859	-	7471B	PPB	11 Feb 2019 15:54:36	U	SMPL
K1423-01 - 1	HB-VJ210	1403	0.2463	-	7471B	PPB	11 Feb 2019 15:56:52	U	SMPL
K1424-01 - 1	SS2-RBR200035	5577	0.9747	-	7471B	PPB	11 Feb 2019 15:59:09	U	SMPL
K1442-04 - 1	COMP-10	7155	1.2501	-	7471B	PPB	11 Feb 2019 16:01:25	U	SMPL
K1442-05 - 1	COMP-11	14223	2.4837	-	7471B	PPB	11 Feb 2019 16:03:42	U	SMPL
K1442-06 - 1	COMP-12	15126	2.6413	-	7471B	PPB	11 Feb 2019 16:05:59	U	SMPL
K1450-01 - 1	SHORE-RD-BOX-CULVERT-TP	12492	2.1816	-	7471B	PPB	11 Feb 2019 16:08:15	U	SMPL
CCV19 - 1	CCV19	27407	4.7846	-	7471B	PPB	11 Feb 2019 16:10:31	U	SMPL
CCB19 - 1	CCB19	68	0.0133	-	7471B	PPB	11 Feb 2019 16:12:47	U	SMPL
K1450-03 - 1	HOSPITAL-BOX-CULVERT-TP	12302	2.1484	-	7471B	PPB	11 Feb 2019 16:15:04	U	SMPL
K1452-01 - 1	TEST-PIT-2	771	0.1360	-	7471B	PPB	11 Feb 2019 16:17:19	U	SMPL
K1453-03 - 1	M-8-S	702	0.1239	-	7471B	PPB	11 Feb 2019 16:19:35	U	SMPL
K1453-07 - 1	M-8-D	711	0.1255	-	7471B	PPB	11 Feb 2019 16:21:51	U	SMPL
K1453-11 - 1	M-9-S	930	0.1637	-	7471B	PPB	11 Feb 2019 16:24:06	U	SMPL
K1453-15 - 1	M-9-D	846	0.1490	-	7471B	PPB	11 Feb 2019 16:26:23	U	SMPL
K1454-01 - 1	TP-1	48958	8.5459	-	7471B	PPB	11 Feb 2019 16:28:39	U	SMPL
K1348-01LX5 - 1		363	0.0647	-	7471B	PPB	11 Feb 2019 16:30:56	U	SMPL
K1348-01A - 1		25854	4.5136	-	7471B	PPB	11 Feb 2019 16:33:15	U	SMPL
CCV20 - 1	CCV20	27934	4.8766	-	7471B	PPB	11 Feb 2019 16:35:31	U	SMPL
CCB20 - 1	CCB20	58	0.0115	-	7471B	PPB	11 Feb 2019 16:37:47	U	SMPL