

OVEN TEMP1: 103 °C TEMP: 105 °C

DATE & TIME IN: 02/11/2019 16:00 OUT: 02/12/2019 08:00

OVEN TEMP2: 105 °C TEMP: 105 °C

DATE & TIME IN: 02/12/2019 09:20 OUT: 02/12/2019 11:00

Run Number: LB100777

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB100777BL	LB100777BL	92.5642	92.5642	100	92.5642	92.5642	92.5642	0.0000	0
2	K1451-01	SHORE-RD-BOX-CULVERT-TP	90.0047	90.0047	100	90.1874	90.1874	90.1874	0.1827	1827
3	K1451-02	HOSPITAL-BOX-CULVERT-TP	90.8002	90.8002	100	91.2547	91.2547	91.2547	0.4545	4545
4	K1451-02DUP	HOSPITAL-BOX-CULVERT-TPDUP	95.0908	95.0908	100	95.5675	95.5675	95.5675	0.4767	4767

A = Sample Volume (ml)
 B = Empty Dish Weight (g)
 C = Final Empty Dish + Sample weight after drying @105°C(g)
 D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS021119

WorkList ID : 122206

Date : 02-11-2019 12:36:29

2B/00777

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/15/2019	Water ✓	K1451-01	TS	Cool 4 deg C	TULL02	L43	SHORE-RD-BOX-CULVERT-T	02/08/2019	SM2540 B
02/15/2019	Water ✓	K1451-02	TS	Cool 4 deg C	TULL02	L43	HOSPITAL-BOX-CULVERT-T	02/08/2019	SM2540 B

Date/Time 02/11/19, 13:00
 Received by: 
 Relinquished by: 

Date/Time 02/11/19, 16:00
 Received by: 
 Relinquished by: 