



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB100782

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.2	4.01	02/11/2019	16:10
2	CAL2	1	Water	NA	NA	22.2	7.00	02/11/2019	16:11
3	CAL3	1	Water	NA	NA	22.2	10.02	02/11/2019	16:12
4	ICV	1	Water	NA	NA	22.1	7.01	02/11/2019	16:14
5	CCV1	1	Water	NA	NA	22.1	2.01	02/11/2019	16:15
6	K1461-09	1	Water	NA	NA	17.6	11.72	02/11/2019	16:18
7	K1461-09DUP	1	Water	NA	NA	17.7	11.73	02/11/2019	16:19
8	CCV2	1	Water	NA	NA	22.1	12.02	02/11/2019	16:20

LB 100782

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K1460

WorkList ID : 122217

Date : 02-11-2019 15:37:46

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/11/2019	Water	K1461-09	pH	Cool 4 deg C	PSEG01	K31	TEST-PIT-4	02/11/2019	9040C

Date/Time 02-11-19 16:00
Received by: Jo
Relinquished by: Co

Date/Time 02-11-19 17:00
Received by: Co
Relinquished by: JP