

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

2/11/2019 16:25

Test: Phenolics- *mg/d*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.027	0.0	0.123	
ICB1	-0.002	0.0	0.016	
CCV1	1.047	0.0	0.125	
CCB1	0.002	0.0	0.017	
PB116982BL	0.002	0.0	0.017	
PB116982BS	1.041	0.0	0.124	
K1451-01	0.018	0.0	0.018	
K1451-01DUP	0.020	0.0	0.019	
K1451-01MS	0.988	0.0	0.118	
K1451-01MSD	0.990	0.0	0.119	
K1451-02	-0.001	0.0	0.016	
CCV2	1.063	0.0	0.126	
CCB2	0.003	0.0	0.017	

N	13
Mean	0.477
SD	0.5296
CV%	111.08

=====

Calibration results	Aquakem 7.2AQ1	Page: 1
---------------------	----------------	---------

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AS

2/11/2019 15:51


Test Phenolics-

Accepted 2/11/2019 15:51

Factor 9.692
 Bias 0.017

Coeff. of det. 0.999923

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.021	0.0456	0.0500	
2	Phenol-2	0.027	0.0969	0.1000	
3	Phenol-2	0.043	0.2571	0.2500	
4	Phenol-2	0.069	0.5074	0.5000	
5	Phenol-2	0.119	0.9912	1.0000	
6	Phenol-2	0.223	2.0020	2.0000	

Aquakem v. 7.2AQ1

Results from time period:

Mon Feb 11 15:51:26 2019

Mon Feb 11 16:19:51 2019

AS

Sample Id	Sam/Ctr/c	Test short name	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0456	mg/l	2/11/2019 15:51:26	
0.1PPM	A	Phenolics-	P	0.0969	mg/l	2/11/2019 15:51:27	
0.25PPM	A	Phenolics-	P	0.2571	mg/l	2/11/2019 15:51:28	
0.50PPM	A	Phenolics-	P	0.5074	mg/l	2/11/2019 15:51:29	
1.0PPM	A	Phenolics-	P	0.9912	mg/l	2/11/2019 15:51:30	
2.0PPM	A	Phenolics-	P	2.002	mg/l	2/11/2019 15:51:31	
ICV1	S	Phenolics-	P	1.0271	mg/l	2/11/2019 16:14:54	
ICB1	S	Phenolics-	P	-0.0023	mg/l	2/11/2019 16:14:55	
CCV1	S	Phenolics-	P	1.0466	mg/l	2/11/2019 16:14:56	
CCB1	S	Phenolics-	P	0.0016	mg/l	2/11/2019 16:14:57	
PB116982BL	S	Phenolics-	P	0.0021	mg/l	2/11/2019 16:14:58	
PB116982BS	S	Phenolics-	P	1.0411	mg/l	2/11/2019 16:14:59	
K1451-01	S	Phenolics-	P	0.0183	mg/l	2/11/2019 16:15:00	
K1451-01DUP	S	Phenolics-	P	0.0204	mg/l	2/11/2019 16:15:01	
K1451-01MS	S	Phenolics-	P	0.9879	mg/l	2/11/2019 16:15:02	
K1451-01MSD	S	Phenolics-	P	0.9902	mg/l	2/11/2019 16:15:03	
K1451-02	S	Phenolics-	P	-0.0008	mg/l	2/11/2019 16:15:04	
CCV2	S	Phenolics-	P	1.0631	mg/l	2/11/2019 16:19:49	
CCB2	S	Phenolics-	P	0.0031	mg/l	2/11/2019 16:19:50	