

LB100789


ISM02.4

INSTRUMENT ID : CV1

Linear

 μ Abs.:

45413



A= 0.0000e+000

B= 2.2028e-004

C= 4.5701e-003

Rho= 0.9998240

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	0.042	0.042	172	0.000	172				
0.2	0.200	0.236	0.036	1049	0.0 %	1049				
2.5	2.500	2.459	-0.041	11142	0.0 %	11142				
5.0	5.000	4.870	-0.130	22089	0.0 %	22089				
7.5	7.500	7.584	0.084	34410	0.0 %	34410				
10.0	10.000	10.008	0.008	45413	0.0 %	45413				

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0	172	-		0 ISM02.4	PPB	2/12/2019 10:14	Std
	0.2	1049	-		0.2 ISM02.4	PPB	2/12/2019 10:17	Std
	2.5	11142	-		2.5 ISM02.4	PPB	2/12/2019 10:19	Std
	5	22089	-		5 ISM02.4	PPB	2/12/2019 10:21	Std
	7.5	34410	-		7.5 ISM02.4	PPB	2/12/2019 10:27	Std
	10	45413	-		10 ISM02.4	PPB	2/12/2019 10:29	Std
ICV	ICV	17632	3.8886 -		ISM02.4	PPB	2/12/2019 10:33	SMPL
ICB	ICB	104	0.0275 -		ISM02.4	PPB	2/12/2019 10:35	SMPL
CCV021	CCV021	22448	4.9494 -		ISM02.4	PPB	2/12/2019 10:38	SMPL
CCB021	CCB021	86	0.0235 -		ISM02.4	PPB	2/12/2019 10:40	SMPL
PB117031BL	PBW031	41	0.0136 -		ISM02.4	PPB	2/12/2019 10:45	SMPL
K1435-06	MH0YM1	221	0.0533 -		ISM02.4	PPB	2/12/2019 10:47	SMPL
K1435-07	MH0YM2	188	0.046 -		ISM02.4	PPB	2/12/2019 10:50	SMPL
PB117048BL	PBW048	4	0.0055 -		ISM02.4	PPB	2/12/2019 10:52	SMPL
K1275-01	MBECX2	185	0.0453 -		ISM02.4	PPB	2/12/2019 10:54	SMPL
K1275-02	MBECX3	187	0.0458 -		ISM02.4	PPB	2/12/2019 10:56	SMPL
K1275-03	MBECX7	254	0.0605 -		ISM02.4	PPB	2/12/2019 10:59	SMPL
K1275-04	MBECX8	237	0.0568 -		ISM02.4	PPB	2/12/2019 11:01	SMPL
K1275-05	MBECX8D	181	0.0444 -		ISM02.4	PPB	2/12/2019 11:03	SMPL
K1275-06	MBECX8S	4802	1.0624 -		ISM02.4	PPB	2/12/2019 11:06	SMPL
K1275-07	MBECX4	147	0.037 -		ISM02.4	PPB	2/12/2019 11:08	SMPL
K1275-08	MBECX6	186	0.0455 -		ISM02.4	PPB	2/12/2019 11:10	SMPL
K1275-09	MBECX9	194	0.0473 -		ISM02.4	PPB	2/12/2019 11:12	SMPL
K1275-10	MBECY0	174	0.0429 -		ISM02.4	PPB	2/12/2019 11:15	SMPL
K1275-11	MBECY1	177	0.0436 -		ISM02.4	PPB	2/12/2019 11:17	SMPL
K1275-12	MBECY2	178	0.0438 -		ISM02.4	PPB	2/12/2019 11:19	SMPL
K1275-13	MBECY3	216	0.0522 -		ISM02.4	PPB	2/12/2019 11:21	SMPL
K1275-14	MBECY5	205	0.0497 -		ISM02.4	PPB	2/12/2019 11:24	SMPL
K1275-15	MBECY6	209	0.0506 -		ISM02.4	PPB	2/12/2019 11:26	SMPL
K1275-16	MBECY8	450	0.1037 -		ISM02.4	PPB	2/12/2019 11:28	SMPL
CCV022	CCV022	22689	5.0025 -		ISM02.4	PPB	2/12/2019 11:31	SMPL
CCB022	CCB022	-41	-0.0045 -		ISM02.4	PPB	2/12/2019 11:33	SMPL
K1275-17	MBECY9	122	0.0314 -		ISM02.4	PPB	2/12/2019 11:35	SMPL
K1275-18	MBECZ0	142	0.0358 -		ISM02.4	PPB	2/12/2019 11:37	SMPL
PB117049BL	PBW049	177	0.0436 -		ISM02.4	PPB	2/12/2019 11:40	SMPL
K1398-01	METC59	151	0.0378 -		ISM02.4	PPB	2/12/2019 11:42	SMPL
K1398-02	METC60	266	0.0632 -		ISM02.4	PPB	2/12/2019 11:44	SMPL
K1398-03	METC60D	233	0.0559 -		ISM02.4	PPB	2/12/2019 11:46	SMPL
K1398-04	METC60S	4941	1.093 -		ISM02.4	PPB	2/12/2019 11:49	SMPL
CCV023	CCV023	22889	5.0466 -		ISM02.4	PPB	2/12/2019 11:51	SMPL
CCB023	CCB023	-55	-0.0075 -		ISM02.4	PPB	2/12/2019 11:53	SMPL