



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB100791

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	02/12/2019	11:15
2	CAL2	1	Water	NA	NA	21.1	7.00	02/12/2019	11:16
3	CAL3	1	Water	NA	NA	21.1	10.02	02/12/2019	11:17
4	ICV	1	Water	NA	NA	21.1	7.01	02/12/2019	11:19
5	CCV1	1	Water	NA	NA	21.1	2.01	02/12/2019	11:20
6	K1467-01	1	Water	NA	NA	17.2	5.27	02/12/2019	11:21
7	K1467-01DUP	1	Water	NA	NA	17.3	5.28	02/12/2019	11:22
8	CCV2	1	Water	NA	NA	21.2	12.01	02/12/2019	11:24

LB100791

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K1467

WorkList ID : 122251

Date : 02-12-2019 11:03:16

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/11/2019	Water	K1467-01	pH	Cool 4 deg C	STER12	L11	CELL-POOL-WATER	02/11/2019	SM 4500-H B

Date/Time 02-12-19 11:10
Received by: JP
Relinquished by: CS

Date/Time 02-12-19 12:00
Received by: CS
Relinquished by: JP