


LB100792
INSTRUMENT ID : CV1

7470A

Linear

μ Abs.:

41423



A= 0.0000e+000

B= 2.4283e-004

C= -4.4529e-002

Rho= 0.9999382

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.004	0.004	200	0.000	200				
0.2	0.200	0.195	-0.005	988	0.0 %	988				
2.5	2.500	2.551	0.051	10690	0.0 %	10690				
5.0	5.000	4.917	-0.083	20432	0.0 %	20432				
7.5	7.500	7.518	0.018	31144	0.0 %	31144				
10.0	10.000	10.014	0.014	41423	0.0 %	41423				

LB100792 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 12 Feb 2019 13:10:58

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	200	-	0.0000	7470A	PPB	12 Feb 2019 13:13:01	S	Std
0.2 - 1	50.2	988	-	0.2000	7470A	PPB	12 Feb 2019 13:15:09	S	Std
2.5 - 1	52.5	10690	-	2.5000	7470A	PPB	12 Feb 2019 13:17:18	S	Std
5.0 - 1	55.0	20432	-	5.0000	7470A	PPB	12 Feb 2019 13:19:25	S	Std
7.5 - 1	57.5	31144	-	7.5000	7470A	PPB	12 Feb 2019 13:21:33	S	Std
10.0 - 1	510.0	41423	-	10.0000	7470A	PPB	12 Feb 2019 13:23:40	S	Std
ICV05 - 1	ICV05	16602	3.9869	-	7470A	PPB	12 Feb 2019 13:26:42	U	SMPL
ICB05 - 1	ICB05	-24	-0.0504	-	7470A	PPB	12 Feb 2019 13:28:49	U	SMPL
CCV24 - 1	CCV24	21560	5.1909	-	7470A	PPB	12 Feb 2019 13:30:54	U	SMPL
CCB24 - 1	CCB24	-53	-0.0574	-	7470A	PPB	12 Feb 2019 13:33:00	U	SMPL
CRA - 1	CRA	991	0.1961	-	7470A	PPB	12 Feb 2019 13:35:06	U	SMPL
HighStd - 1	HighStd	41089	9.9331	-	7470A	PPB	12 Feb 2019 13:37:12	U	SMPL
ChkStd - 1	ChkStd	28778	6.9436	-	7470A	PPB	12 Feb 2019 13:39:18	U	SMPL
PB117028BL - 1	PBW	115	-0.0166	-	7470A	PPB	12 Feb 2019 13:41:26	U	SMPL
PB117028BS - 1	LCSW	16559	3.9765	-	7470A	PPB	12 Feb 2019 13:43:33	U	SMPL
K1356-06 - 1	SU-03-020819-A	428	0.0594	-	7470A	PPB	12 Feb 2019 13:45:40	U	SMPL
K1356-06DUP - 1	SU-03-020819-ADUP	431	0.0601	-	7470A	PPB	12 Feb 2019 13:47:47	U	SMPL
K1356-06MS - 1	SU-03-020819-AMS	16737	4.0197	-	7470A	PPB	12 Feb 2019 13:49:55	U	SMPL
K1356-06MSD - 1	SU-03-020819-AMSD	16650	3.9986	-	7470A	PPB	12 Feb 2019 13:52:03	U	SMPL
K1356-08 - 1	SU-03-020819-B	197	0.0033	-	7470A	PPB	12 Feb 2019 13:54:09	U	SMPL
CCV25 - 1	CCV25	21404	5.1530	-	7470A	PPB	12 Feb 2019 13:56:15	U	SMPL
CCB25 - 1	CCB25	-35	-0.0530	-	7470A	PPB	12 Feb 2019 13:58:21	U	SMPL
K1423-03 - 1	HB-VJ210	527	0.0834	-	7470A	PPB	12 Feb 2019 14:00:27	U	SMPL
K1424-03 - 1	SS2-RBR200035	533	0.0849	-	7470A	PPB	12 Feb 2019 14:02:34	U	SMPL
K1442-07 - 1	COMP-10	571	0.0941	-	7470A	PPB	12 Feb 2019 14:04:40	U	SMPL
K1442-08 - 1	COMP-11	188	0.0011	-	7470A	PPB	12 Feb 2019 14:06:47	U	SMPL
K1442-09 - 1	COMP-12	208	0.0060	-	7470A	PPB	12 Feb 2019 14:08:53	U	SMPL
K1450-02 - 1	SHORE-RD-BOX-CULVERT-TP	183	-0.0001	-	7470A	PPB	12 Feb 2019 14:11:00	U	SMPL
K1450-04 - 1	HOSPITAL-BOX-CULVERT-TP	183	-0.0001	-	7470A	PPB	12 Feb 2019 14:13:08	U	SMPL
K1452-04 - 1	TEST-PIT-2	179	-0.0011	-	7470A	PPB	12 Feb 2019 14:15:15	U	SMPL
K1453-04 - 1	M-8-S	183	-0.0001	-	7470A	PPB	12 Feb 2019 14:17:23	U	SMPL
K1453-08 - 1	M-8-D	165	-0.0045	-	7470A	PPB	12 Feb 2019 14:19:30	U	SMPL
CCV26 - 1	CCV26	20377	4.9036	-	7470A	PPB	12 Feb 2019 14:21:36	U	SMPL
CCB26 - 1	CCB26	-50	-0.0567	-	7470A	PPB	12 Feb 2019 14:23:42	U	SMPL
K1453-12 - 1	M-9-S	89	-0.0229	-	7470A	PPB	12 Feb 2019 14:25:48	U	SMPL
K1453-16 - 1	M-9-D	84	-0.0241	-	7470A	PPB	12 Feb 2019 14:27:55	U	SMPL
K1454-04 - 1	TP-1	102	-0.0198	-	7470A	PPB	12 Feb 2019 14:30:01	U	SMPL
PB117029BL - 1	PBW	146	-0.0091	-	7470A	PPB	12 Feb 2019 14:32:08	U	SMPL
PB117029BS - 1	LCSW	16035	3.8492	-	7470A	PPB	12 Feb 2019 14:34:15	U	SMPL
K1425-01 - 1	PM-1	152	-0.0076	-	7470A	PPB	12 Feb 2019 14:36:22	U	SMPL
K1425-02 - 1	PM-2	187	0.0009	-	7470A	PPB	12 Feb 2019 14:38:29	U	SMPL
K1425-03 - 1	PM-3	214	0.0074	-	7470A	PPB	12 Feb 2019 14:40:37	U	SMPL
K1425-04 - 1	PM-4	212	0.0070	-	7470A	PPB	12 Feb 2019 14:42:45	U	SMPL
K1425-05 - 1	PM-5	184	0.0002	-	7470A	PPB	12 Feb 2019 14:44:51	U	SMPL
CCV27 - 1	CCV27	20495	4.9323	-	7470A	PPB	12 Feb 2019 14:46:58	U	SMPL
CCB27 - 1	CCB27	-52	-0.0572	-	7470A	PPB	12 Feb 2019 14:49:04	U	SMPL
K1425-06 - 1	PM-6	185	0.0004	-	7470A	PPB	12 Feb 2019 14:51:11	U	SMPL
K1425-07 - 1	PM-7	183	-0.0001	-	7470A	PPB	12 Feb 2019 14:53:17	U	SMPL
K1425-08 - 1	PM-8	178	-0.0013	-	7470A	PPB	12 Feb 2019 14:55:24	U	SMPL
K1425-09 - 1	PM-9	182	-0.0003	-	7470A	PPB	12 Feb 2019 14:57:30	U	SMPL
K1425-10 - 1	PM-10	190	0.0016	-	7470A	PPB	12 Feb 2019 14:59:37	U	SMPL
K1425-11 - 1	PM-11	140	-0.0105	-	7470A	PPB	12 Feb 2019 15:01:44	U	SMPL
K1425-12 - 1	PM-12	150	-0.0081	-	7470A	PPB	12 Feb 2019 15:03:52	U	SMPL
K1425-13 - 1	PM-13	197	0.0033	-	7470A	PPB	12 Feb 2019 15:06:00	U	SMPL
K1425-14 - 1	PM-14	175	-0.0020	-	7470A	PPB	12 Feb 2019 15:08:08	U	SMPL
K1425-15 - 1	PM-15	154	-0.0071	-	7470A	PPB	12 Feb 2019 15:10:14	U	SMPL
CCV28 - 1	CCV28	20943	5.0410	-	7470A	PPB	12 Feb 2019 15:12:21	U	SMPL
CCB28 - 1	CCB28	-28	-0.0513	-	7470A	PPB	12 Feb 2019 15:14:27	U	SMPL
K1425-16 - 1	PM-16	156	-0.0066	-	7470A	PPB	12 Feb 2019 15:16:33	U	SMPL
K1425-17 - 1	PM-17	194	0.0026	-	7470A	PPB	12 Feb 2019 15:18:40	U	SMPL
K1425-18 - 1	PM-18	168	-0.0037	-	7470A	PPB	12 Feb 2019 15:20:46	U	SMPL
K1425-19 - 1	PM-19	179	-0.0011	-	7470A	PPB	12 Feb 2019 15:22:53	U	SMPL
K1425-20 - 1	PM-20	171	-0.0030	-	7470A	PPB	12 Feb 2019 15:25:00	U	SMPL
K1425-20DUP - 1	PM-20DUP	168	-0.0037	-	7470A	PPB	12 Feb 2019 15:27:08	U	SMPL
K1425-20MS - 1	PM-20MS	16553	3.9750	-	7470A	PPB	12 Feb 2019 15:29:15	U	SMPL
K1425-20MSD - 1	PM-20MSD	16257	3.9031	-	7470A	PPB	12 Feb 2019 15:31:23	U	SMPL
PB117030BL - 1	PBW	32	-0.0368	-	7470A	PPB	12 Feb 2019 15:33:31	U	SMPL
PB117030BS - 1	LCSW	15741	3.7778	-	7470A	PPB	12 Feb 2019 15:35:38	U	SMPL
CCV29 - 1	CCV29	21462	5.1671	-	7470A	PPB	12 Feb 2019 15:37:45	U	SMPL
CCB29 - 1	CCB29	-61	-0.0593	-	7470A	PPB	12 Feb 2019 15:39:53	U	SMPL
K1405-01 - 1	26-WARD-UNK-LIQUID	2244	0.5004	-	7470A	PPB	12 Feb 2019 15:42:00	U	SMPL
K1406-01 - 1	N.RIVER-UNK-LIQUID	-316	-0.1213	-	7470A	PPB	12 Feb 2019 15:44:07	U	SMPL
K1448-01 - 1	30171	-1	-0.0448	-	7470A	PPB	12 Feb 2019 15:46:16	U	SMPL
K1448-02 - 1	30172	82	-0.0246	-	7470A	PPB	12 Feb 2019 15:48:23	U	SMPL
K1448-02DUP - 1	30172DUP	95	-0.0215	-	7470A	PPB	12 Feb 2019 15:50:31	U	SMPL
K1448-02MS - 1	30172MS	18337	4.4082	-	7470A	PPB	12 Feb 2019 15:52:39	U	SMPL
K1448-02MSD - 1	30172MSD	18228	4.3818	-	7470A	PPB	12 Feb 2019 15:54:46	U	SMPL
K1448-03 - 1	30228	93	-0.0219	-	7470A	PPB	12 Feb 2019 15:56:54	U	SMPL
K1448-04 - 1	30226	65	-0.0287	-	7470A	PPB	12 Feb 2019 15:59:02	U	SMPL
PB117028TB - 1	PB116967TB	12	-0.0416	-	7470A	PPB	12 Feb 2019 16:04:45	U	SMPL
CCV30 - 1	CCV30	21156	5.0928	-	7470A	PPB	12 Feb 2019 16:06:52	U	SMPL

LB100792 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 12 Feb 2019 13:10:58

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB30 - 1	CCB30	136	-0.0115		-7470A	PPB	12 Feb 2019 16:08:59	U	SMPL
PB117029TB - 1	PB116983TB	8	-0.0426		-7470A	PPB	12 Feb 2019 16:11:06	U	SMPL
PB117030TB - 1	PB116969TB	5	-0.0433		-7470A	PPB	12 Feb 2019 16:13:13	U	SMPL
K1356-06LX5 - 1		102	-0.0198		-7470A	PPB	12 Feb 2019 16:15:20	U	SMPL
K1356-06A - 1		18083	4.3465		-7470A	PPB	12 Feb 2019 16:17:27	U	SMPL
K1425-20LX5 - 1		-6	-0.0460		-7470A	PPB	12 Feb 2019 16:19:34	U	SMPL
K1425-20A - 1		17378	4.1754		-7470A	PPB	12 Feb 2019 16:21:42	U	SMPL
K1448-02LX5 - 1		-14	-0.0479		-7470A	PPB	12 Feb 2019 16:23:49	U	SMPL
K1448-02A - 1		19247	4.6292		-7470A	PPB	12 Feb 2019 16:25:57	U	SMPL
CCV31 - 1	CCV31	20548	4.9451		-7470A	PPB	12 Feb 2019 16:28:05	U	SMPL
CCB31 - 1	CCB31	41	-0.0346		-7470A	PPB	12 Feb 2019 16:30:12	U	SMPL