

LB100812


INSTRUMENT ID: CV1

245.1

Linear

 μ Abs.:

38756



A= 0.0000e+000

B= 2.5543e-004

C= -7.1718e-002

Rho= 0.9995624

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.076	-0.076	-15	0.000	-15				
0.05	0.050	0.063	0.013	526	0.0 %	526				
0.2	0.200	0.194	-0.006	1040	0.0 %	1040				
2.5	2.500	2.445	-0.055	9852	0.0 %	9852				
5.0	5.000	5.179	0.179	20555	0.0 %	20555				
7.5	7.500	7.618	0.118	30104	0.0 %	30104				
10.0	10.000	9.828	-0.172	38756	0.0 %	38756				

LB100812 B INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 13 Feb 2019 12:55:59

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0-0	-15	-	0.0000	245.1	PPB	13 Feb 2019 13:00:58	S	Std
0.05 - 1	S0-05	526	-	0.0500	245.1	PPB	13 Feb 2019 13:03:16	S	Std
0.2 - 1	S0-2	1040	-	0.2000	245.1	PPB	13 Feb 2019 13:05:33	S	Std
2.5 - 1	S2-5	9852	-	2.5000	245.1	PPB	13 Feb 2019 13:07:50	S	Std
5.0 - 1	S5-0	20555	-	5.0000	245.1	PPB	13 Feb 2019 13:10:07	S	Std
7.5 - 1	S7-5	30104	-	7.5000	245.1	PPB	13 Feb 2019 13:26:26	S	Std
10.0 - 1	S10-0	38756	-	10.0000	245.1	PPB	13 Feb 2019 13:28:42	S	Std
ICV07 - 1	ICV07	15493	3.8857	-245.1		PPB	13 Feb 2019 13:32:30	U	SMPL
ICB07 - 1	ICB07	-66	-0.0886	-245.1		PPB	13 Feb 2019 13:34:45	U	SMPL
CCV39 - 1	CCV39	19877	5.0055	-245.1		PPB	13 Feb 2019 13:37:01	U	SMPL
CCB39 - 1	CCB39	-11	-0.0745	-245.1		PPB	13 Feb 2019 13:39:16	U	SMPL
CRA - 1	CRA	528	0.0631	-245.1		PPB	13 Feb 2019 13:49:05	U	SMPL
HighStd - 1	HighStd	39704	10.0699	-245.1		PPB	13 Feb 2019 13:51:20	U	SMPL
ChkStd - 1	ChkStd	27709	7.0060	-245.1		PPB	13 Feb 2019 13:53:36	U	SMPL
PB117026BL - 1	PBW	110	-0.0436	-245.1		PPB	13 Feb 2019 13:55:53	U	SMPL
PB117026BS - 1	LCSW	16170	4.0586	-245.1		PPB	13 Feb 2019 13:58:11	U	SMPL
K1402-01 - 1	COMP	235	-0.0117	-245.1		PPB	13 Feb 2019 14:00:27	U	SMPL
K1402-01DUP - 1	COMPDUP	255	-0.0066	-245.1		PPB	13 Feb 2019 14:02:44	U	SMPL
K1402-01MS - 1	COMPMS	16703	4.1947	-245.1		PPB	13 Feb 2019 14:05:01	U	SMPL
K1402-01MSD - 1	COMPMSD	16656	4.1827	-245.1		PPB	13 Feb 2019 14:07:18	U	SMPL
K1421-06 - 1	HP-640	280	-0.0002	-245.1		PPB	13 Feb 2019 14:09:34	U	SMPL
CCV40 - 1	CCV40	19583	4.9304	-245.1		PPB	13 Feb 2019 14:11:49	U	SMPL
CCB40 - 1	CCB40	34	-0.0630	-245.1		PPB	13 Feb 2019 14:14:05	U	SMPL
K1451-01 - 1	SHORE-RD-BOX-CULVERT-TP	18752	4.7181	-245.1		PPB	13 Feb 2019 14:16:20	U	SMPL
K1451-02 - 1	HOSPITAL-BOX-CULVERT-TP	1883	0.4093	-245.1		PPB	13 Feb 2019 14:18:36	U	SMPL
K1402-01LX5 - 1		37	-0.0623	-245.1		PPB	13 Feb 2019 14:20:51	U	SMPL
K1402-01A - 1		18509	4.6561	-245.1		PPB	13 Feb 2019 14:23:07	U	SMPL
CCV41 - 1	CCV41	19943	5.0223	-245.1		PPB	13 Feb 2019 14:25:23	U	SMPL
CCB41 - 1	CCB41	21	-0.0664	-245.1		PPB	13 Feb 2019 14:27:39	U	SMPL