



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB100818

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	6	27.0	80.6	80.9	02/13/2019	09:10
2	K1460-09		1	8	>100.0	>212.0	>212.0	02/13/2019	09:40
3	K1460-09DUP		1	7	>100.0	>212.0	>212.0	02/13/2019	10:10
4	K1461-10		1	9	>100.0	>212.0	>212.0	02/13/2019	10:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB100818

Worklist Name : FP21219

Worklist ID : 122229

Date : 02-11-2019 16:26:56

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/22/2019	Water	K1460-09	Flash Point	Cool 4 deg C	CAMP02		WC-20190208	02/08/2019	1010A
02/14/2019	Water	K1461-10	Flash Point	Cool 4 deg C	PSEG01	K31	TEST-PIT-4	02/11/2019	1010A

Date/Time 02/13/19 8:30
 Received by: KS
 Relinquished by: CP

Date/Time 02/13/19 10:55
 Received by: CP
 Relinquished by: KS