

LB100820


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

44954



A= 0.0000e+000

B= 2.2583e-004

C= -6.4863e-002

Rho= 0.9997698

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.004	-0.004	268	0.000	268				
0.2	0.200	0.188	-0.012	1120	0.0 %	1120				
2.5	2.500	2.612	0.112	11853	0.0 %	11853				
5.0	5.000	4.926	-0.074	22101	0.0 %	22101				
7.5	7.500	7.391	-0.109	33015	0.0 %	33015				
10.0	10.000	10.087	0.087	44954	0.0 %	44954				

LB100820 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 Feb 2019 15:24:28

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	268	-	0.0000	7470A	PPB	13 Feb 2019 15:26:57	S	Std
0.2 - 1	50-2	1120	-	0.2000	7470A	PPB	13 Feb 2019 15:29:05	S	Std
2.5 - 1	52-5	11853	-	2.5000	7470A	PPB	13 Feb 2019 15:31:13	S	Std
5.0 - 1	55-0	22101	-	5.0000	7470A	PPB	13 Feb 2019 15:33:21	S	Std
7.5 - 1	57-5	33015	-	7.5000	7470A	PPB	13 Feb 2019 15:35:28	S	Std
10.0 - 1	510-0	44954	-	10.0000	7470A	PPB	13 Feb 2019 15:37:35	S	Std
ICV08 - 1	ICV08	18604	4.1365	-	7470A	PPB	13 Feb 2019 15:41:38	U	SMPL
ICB08 - 1	ICB08	41	-0.0556	-	7470A	PPB	13 Feb 2019 15:43:44	U	SMPL
CCV42 - 1	CCV42	22173	4.9425	-	7470A	PPB	13 Feb 2019 15:45:50	U	SMPL
CCB42 - 1	CCB42	5	-0.0637	-	7470A	PPB	13 Feb 2019 15:47:56	U	SMPL
CRA - 1	CRA	1136	0.1917	-	7470A	PPB	13 Feb 2019 15:50:02	U	SMPL
HighStd - 1	HighStd	44998	10.0971	-	7470A	PPB	13 Feb 2019 15:52:08	U	SMPL
ChkStd - 1	ChkStd	29934	6.6952	-	7470A	PPB	13 Feb 2019 15:54:14	U	SMPL
PB117027BL - 1	PBW	107	-0.0407	-	7470A	PPB	13 Feb 2019 15:56:22	U	SMPL
PB117027BS - 1	LCSW	15724	3.4861	-	7470A	PPB	13 Feb 2019 15:58:30	U	SMPL
K1449-01 - 1	B-19-OW	1435	0.2592	-	7470A	PPB	13 Feb 2019 16:00:37	U	SMPL
K1449-02 - 1	B-23-OW	1345	0.2389	-	7470A	PPB	13 Feb 2019 16:02:44	U	SMPL
K1449-03 - 1	B-6-OW	8497	1.8540	-	7470A	PPB	13 Feb 2019 16:04:52	U	SMPL
K1449-04 - 1	B-17-OW	1782	0.3376	-	7470A	PPB	13 Feb 2019 16:07:00	U	SMPL
K1449-05 - 1	B-11-OW	54782	12.3066	-	7470A	PPB	13 Feb 2019 16:09:06	U	SMPL
CCV43 - 1	CCV43	21912	4.8836	-	7470A	PPB	13 Feb 2019 16:11:12	U	SMPL
CCB43 - 1	CCB43	-53	-0.0768	-	7470A	PPB	13 Feb 2019 16:13:22	U	SMPL
K1449-06 - 1	B-19-OW	249	-0.0086	-	7470A	PPB	13 Feb 2019 16:15:28	U	SMPL
K1449-06DUP - 1	B-19-OWDUP	263	-0.0055	-	7470A	PPB	13 Feb 2019 16:17:34	U	SMPL
K1449-06MS - 1	B-19-OWMS	16655	3.6964	-	7470A	PPB	13 Feb 2019 16:19:40	U	SMPL
K1449-06MSD - 1	B-19-OWMSD	16681	3.7022	-	7470A	PPB	13 Feb 2019 16:21:46	U	SMPL
K1449-07 - 1	B-23-OW	214	-0.0165	-	7470A	PPB	13 Feb 2019 16:23:52	U	SMPL
K1449-08 - 1	B-6-OW	226	-0.0138	-	7470A	PPB	13 Feb 2019 16:26:00	U	SMPL
K1449-09 - 1	B-17-OW	278	-0.0021	-	7470A	PPB	13 Feb 2019 16:28:07	U	SMPL
K1449-10 - 1	B-11-OW	284	-0.0007	-	7470A	PPB	13 Feb 2019 16:30:16	U	SMPL
PB117069BL - 1	PBW	154	-0.0301	-	7470A	PPB	13 Feb 2019 16:32:25	U	SMPL
PB117069BS - 1	LCSW	16020	3.5530	-	7470A	PPB	13 Feb 2019 16:34:31	U	SMPL
CCV44 - 1	CCV44	21963	4.8951	-	7470A	PPB	13 Feb 2019 16:36:38	U	SMPL
CCB44 - 1	CCB44	-19	-0.0692	-	7470A	PPB	13 Feb 2019 16:38:45	U	SMPL
K1457-07 - 1	IDW-LIQUID-01	137	-0.0339	-	7470A	PPB	13 Feb 2019 16:40:51	U	SMPL
K1457-07DUP - 1	IDW-LIQUID-01DUP	149	-0.0312	-	7470A	PPB	13 Feb 2019 16:42:57	U	SMPL
K1457-07MS - 1	IDW-LIQUID-01MS	16790	3.7268	-	7470A	PPB	13 Feb 2019 16:45:04	U	SMPL
K1457-07MSD - 1	IDW-LIQUID-01MSD	16769	3.7221	-	7470A	PPB	13 Feb 2019 16:47:10	U	SMPL
K1457-08 - 1	IDW-LIQUID-02	91	-0.0443	-	7470A	PPB	13 Feb 2019 16:49:17	U	SMPL
K1457-09 - 1	IDW-LIQUID-03	148	-0.0314	-	7470A	PPB	13 Feb 2019 16:51:25	U	SMPL
K1457-10 - 1	IDW-LIQUID-04	149	-0.0312	-	7470A	PPB	13 Feb 2019 16:53:32	U	SMPL
K1460-09 - 1	WC-20190208	136	-0.0342	-	7470A	PPB	13 Feb 2019 16:55:40	U	SMPL
K1461-10 - 1	TEST-PIT-4	150	-0.0310	-	7470A	PPB	13 Feb 2019 16:57:47	U	SMPL
PB117070BL - 1	PBW	144	-0.0323	-	7470A	PPB	13 Feb 2019 16:59:54	U	SMPL
CCV45 - 1	CCV45	22047	4.9140	-	7470A	PPB	13 Feb 2019 17:02:00	U	SMPL
CCB45 - 1	CCB45	16	-0.0612	-	7470A	PPB	13 Feb 2019 17:04:07	U	SMPL
PB117070BS - 1	LCSW	15514	3.4387	-	7470A	PPB	13 Feb 2019 17:06:13	U	SMPL
K1345-02 - 1	HD-02-021119	117	-0.0384	-	7470A	PPB	13 Feb 2019 17:08:20	U	SMPL
K1345-02DUP - 1	HD-02-021119DUP	136	-0.0342	-	7470A	PPB	13 Feb 2019 17:10:26	U	SMPL
K1345-02MS - 1	HD-02-021119MS	16479	3.6566	-	7470A	PPB	13 Feb 2019 17:12:33	U	SMPL
K1345-02MSD - 1	HD-02-021119MSD	16566	3.6763	-	7470A	PPB	13 Feb 2019 17:14:40	U	SMPL
K1457-01 - 1	IDW-SOIL-01	101	-0.0421	-	7470A	PPB	13 Feb 2019 17:16:47	U	SMPL
K1457-02 - 1	IDW-SOIL-02	151	-0.0308	-	7470A	PPB	13 Feb 2019 17:18:55	U	SMPL
K1457-03 - 1	IDW-SOIL-03	126	-0.0364	-	7470A	PPB	13 Feb 2019 17:21:02	U	SMPL
K1457-04 - 1	IDW-SOIL-04	150	-0.0310	-	7470A	PPB	13 Feb 2019 17:23:10	U	SMPL
K1458-02 - 1	WC-1	133	-0.0348	-	7470A	PPB	13 Feb 2019 17:25:16	U	SMPL
CCV46 - 1	CCV46	22412	4.9965	-	7470A	PPB	13 Feb 2019 17:27:23	U	SMPL
CCB46 - 1	CCB46	-7	-0.0664	-	7470A	PPB	13 Feb 2019 17:29:29	U	SMPL
K1458-07 - 1	WC-4	150	-0.0310	-	7470A	PPB	13 Feb 2019 17:31:36	U	SMPL
K1458-10 - 1	WC-6	159	-0.0290	-	7470A	PPB	13 Feb 2019 17:33:42	U	SMPL
K1459-04 - 1	M-10-S	162	-0.0283	-	7470A	PPB	13 Feb 2019 17:35:48	U	SMPL
K1459-08 - 1	M-10-D	162	-0.0283	-	7470A	PPB	13 Feb 2019 17:37:55	U	SMPL
K1459-12 - 1	M-11-S	158	-0.0292	-	7470A	PPB	13 Feb 2019 17:40:02	U	SMPL
K1461-04 - 1	TEST-PIT-3	158	-0.0292	-	7470A	PPB	13 Feb 2019 17:42:09	U	SMPL
K1461-08 - 1	TEST-PIT-4	154	-0.0301	-	7470A	PPB	13 Feb 2019 17:44:16	U	SMPL
K1462-04 - 1	TP-1-S	124	-0.0369	-	7470A	PPB	13 Feb 2019 17:46:24	U	SMPL
K1462-08 - 1	TP-1-D	182	-0.0238	-	7470A	PPB	13 Feb 2019 17:48:32	U	SMPL
K1462-12 - 1	TP-2-S	154	-0.0301	-	7470A	PPB	13 Feb 2019 17:50:39	U	SMPL
CCV47 - 1	CCV47	22519	5.0206	-	7470A	PPB	13 Feb 2019 17:52:47	U	SMPL
CCB47 - 1	CCB47	-3	-0.0655	-	7470A	PPB	13 Feb 2019 17:54:53	U	SMPL
K1462-16 - 1	TP-2-D	140	-0.0332	-	7470A	PPB	13 Feb 2019 17:57:01	U	SMPL
K1462-20 - 1	TP-3-S	227	-0.0136	-	7470A	PPB	13 Feb 2019 17:59:08	U	SMPL
K1462-24 - 1	TP-3-D	160	-0.0287	-	7470A	PPB	13 Feb 2019 18:01:15	U	SMPL
PB117071BL - 1	PBW	132	-0.0351	-	7470A	PPB	13 Feb 2019 18:03:22	U	SMPL
PB117071BS - 1	LCSW	17642	3.9193	-	7470A	PPB	13 Feb 2019 18:05:29	U	SMPL
K1404-01 - 1	FB-02-06-19	191	-0.0217	-	7470A	PPB	13 Feb 2019 18:07:36	U	SMPL
K1460-01 - 1	WC-20190208	427	0.0316	-	7470A	PPB	13 Feb 2019 18:09:44	U	SMPL
K1460-01DUP - 1	WC-20190208DUP	423	0.0307	-	7470A	PPB	13 Feb 2019 18:11:52	U	SMPL
K1460-01MS - 1	WC-20190208MS	18294	4.0665	-	7470A	PPB	13 Feb 2019 18:14:00	U	SMPL
K1460-01MSD - 1	WC-20190208MSD	18195	4.0441	-	7470A	PPB	13 Feb 2019 18:16:07	U	SMPL
CCV48 - 1	CCV48	20524	4.5701	-	7470A	PPB	13 Feb 2019 18:18:14	U	SMPL

LB100820 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 Feb 2019 15:24:28

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB48 - 1	CCB48	-21	-0.0696		-7470A	PPB	13 Feb 2019 18:20:21	U	SMPL
K1461-09 - 1	TEST-PIT-4	5571	1.1932		-7470A	PPB	13 Feb 2019 18:22:28	U	SMPL
PB117069TB - 1	PB117024TB	102	-0.0418		-7470A	PPB	13 Feb 2019 18:24:36	U	SMPL
PB117070TB - 1	PB117022TB	117	-0.0384		-7470A	PPB	13 Feb 2019 18:26:43	U	SMPL
K1449-06LX5 - 1		41	-0.0556		-7470A	PPB	13 Feb 2019 18:28:50	U	SMPL
K1449-06A - 1		18458	4.1035		-7470A	PPB	13 Feb 2019 18:30:57	U	SMPL
K1457-07LX5 - 1		2	-0.0644		-7470A	PPB	13 Feb 2019 18:33:04	U	SMPL
K1457-07A - 1		16876	3.7463		-7470A	PPB	13 Feb 2019 18:35:11	U	SMPL
K1345-02LX5 - 1		-13	-0.0678		-7470A	PPB	13 Feb 2019 18:37:19	U	SMPL
K1345-02A - 1		16725	3.7122		-7470A	PPB	13 Feb 2019 18:39:27	U	SMPL
CCV49 - 1	CCV49	22428	5.0001		-7470A	PPB	13 Feb 2019 18:41:34	U	SMPL
CCB49 - 1	CCB49	-31	-0.0719		-7470A	PPB	13 Feb 2019 18:43:41	U	SMPL
K1460-01LX5 - 1		84	-0.0459		-7470A	PPB	13 Feb 2019 18:45:48	U	SMPL
K1460-01A - 1		17428	3.8709		-7470A	PPB	13 Feb 2019 18:47:55	U	SMPL
K1449-05DLX2 - 1		27171	6.0712		-7470A	PPB	13 Feb 2019 18:50:02	U	SMPL
CCV50 - 1	CCV50	21132	4.7074		-7470A	PPB	13 Feb 2019 18:52:09	U	SMPL
CCB50 - 1	CCB50	-61	-0.0786		-7470A	PPB	13 Feb 2019 18:54:18	U	SMPL