



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB100821

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.1	4.01	02/13/2019	16:50
2	CAL2	1	Water	NA	NA	22.1	7.01	02/13/2019	16:51
3	CAL3	1	Water	NA	NA	22.1	10.03	02/13/2019	16:52
4	ICV	1	Water	NA	NA	22.1	6.99	02/13/2019	17:00
5	CCV1	1	Water	NA	NA	22.1	2.01	02/13/2019	17:01
6	K1349-06	1	Solid	20.03	20	21.4	8.04	02/13/2019	17:02
7	K1349-06DUP	1	Solid	20.04	20	21.5	8.05	02/13/2019	17:03
8	K1349-08	1	Solid	20.05	20	20.8	7.87	02/13/2019	17:05
9	K1350-06	1	Solid	20.02	20	20.4	8.75	02/13/2019	17:06
10	K1350-08	1	Solid	20.03	20	20.2	8.67	02/13/2019	17:08
11	K1350-10	1	Solid	20.04	20	20.4	8.58	02/13/2019	17:09
12	K1479-01	1	Solid	20.02	20	18.9	6.74	02/13/2019	17:10
13	K1480-01	1	Solid	20.01	20	19.0	6.46	02/13/2019	17:11
14	CCV2	1	Water	NA	NA	22.2	12.02	02/13/2019	17:12

LB100821

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K149

WorkList ID : 122294

Date : 02-13-2019 14:58:42

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/13/2019	Solid	K1349-06	Corrosivity	Cool 4 deg C	PSEG05		OR-01-021319-A	02/13/2019	9045D
02/13/2019	Solid	K1349-08	Corrosivity	Cool 4 deg C	PSEG05		OR-01-021319-B	02/13/2019	9045D
02/13/2019	Solid	K1350-06	Corrosivity	Cool 4 deg C	PSEG05		EO-02-021319-A	02/13/2019	9045D
02/13/2019	Solid	K1350-08	Corrosivity	Cool 4 deg C	PSEG05		EO-02-021319-B	02/13/2019	9045D
02/13/2019	Solid	K1350-10	Corrosivity	Cool 4 deg C	PSEG05		EO-02-021319-C	02/13/2019	9045D
02/11/2019	Solid	K1479-01	Corrosivity	Cool 4 deg C	CLEA11	L11	OIL-WATER-SEPERATOR	02/11/2019	9045D
02/12/2019	Solid	K1480-01	Corrosivity	Cool 4 deg C	CLEA11	L12	OIL-WATER-SEPERATOR-3	02/12/2019	9045D

Date/Time 02-13-19 16:10
 Received by: JD
 Relinquished by: SC

Date/Time 02-13-19 17:35
 Received by: JD
 Relinquished by: SC