

LB100827


INSTRUMENT ID: CV1

7470A

Linear

μ Abs.:

40121



A= 0.0000e+000

B= 2.4883e-004

C= -1.2161e-002

Rho= 0.9999845

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.000	0.000	50	0.000	50				
0.2	0.200	0.174	-0.026	748	0.0 %	748				
2.5	2.500	2.515	0.015	10155	0.0 %	10155				
5.0	5.000	5.018	0.018	20215	0.0 %	20215				
7.5	7.500	7.522	0.022	30279	0.0 %	30279				
10.0	10.000	9.971	-0.029	40121	0.0 %	40121				

LB100827 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 14 Feb 2019 08:28:55

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	50	-	0.0000	7470A	PPB	14 Feb 2019 08:56:57	S	Std
0.2 - 1	50.2	748	-	0.2000	7470A	PPB	14 Feb 2019 08:59:07	S	Std
2.5 - 1	52.5	10155	-	2.5000	7470A	PPB	14 Feb 2019 09:01:16	S	Std
5.0 - 1	55.0	20215	-	5.0000	7470A	PPB	14 Feb 2019 09:03:24	S	Std
7.5 - 1	57.5	30279	-	7.5000	7470A	PPB	14 Feb 2019 09:05:32	S	Std
10.0 - 1	50.0	40121	-	10.0000	7470A	PPB	14 Feb 2019 09:07:40	S	Std
ICV09 - 1	ICV09	15739	3.9041	-	7470A	PPB	14 Feb 2019 09:17:15	U	SMPL
ICB09 - 1	ICB09	-33	-0.0204	-	7470A	PPB	14 Feb 2019 09:19:21	U	SMPL
CCV51 - 1	CCV51	19872	4.9325	-	7470A	PPB	14 Feb 2019 09:21:27	U	SMPL
CCB51 - 1	CCB51	-28	-0.0191	-	7470A	PPB	14 Feb 2019 09:23:33	U	SMPL
CRA - 1	CRA	883	0.2076	-	7470A	PPB	14 Feb 2019 09:25:39	U	SMPL
HighStd - 1	HighStd	39599	9.8412	-	7470A	PPB	14 Feb 2019 09:27:45	U	SMPL
ChkStd - 1	ChkStd	26900	6.6813	-	7470A	PPB	14 Feb 2019 09:29:51	U	SMPL
PB117098BL - 1	PBW	70	0.0053	-	7470A	PPB	14 Feb 2019 09:31:59	U	SMPL
PB117098BS - 1	LCSW	16655	4.1321	-	7470A	PPB	14 Feb 2019 09:34:07	U	SMPL
K1343-02 - 1	JC-02-021119-A	157	0.0269	-	7470A	PPB	14 Feb 2019 09:36:14	U	SMPL
K1343-02DUP - 1	JC-02-021119-ADUP	172	0.0306	-	7470A	PPB	14 Feb 2019 09:38:21	U	SMPL
K1343-02MS - 1	JC-02-021119-AMS	16952	4.2060	-	7470A	PPB	14 Feb 2019 09:40:28	U	SMPL
K1343-02MSD - 1	JC-02-021119-AMSD	16653	4.1316	-	7470A	PPB	14 Feb 2019 09:42:37	U	SMPL
K1343-04 - 1	JC-02-021119-B	205	0.0388	-	7470A	PPB	14 Feb 2019 09:44:43	U	SMPL
CCV52 - 1	CCV52	19968	4.9564	-	7470A	PPB	14 Feb 2019 09:46:49	U	SMPL
CCB52 - 1	CCB52	-45	-0.0234	-	7470A	PPB	14 Feb 2019 09:48:56	U	SMPL
K1343-06 - 1	JC-02-021119-C	220	0.0426	-	7470A	PPB	14 Feb 2019 09:51:02	U	SMPL
K1343-08 - 1	JC-02-021119-D	229	0.0448	-	7470A	PPB	14 Feb 2019 09:53:08	U	SMPL
K1343-10 - 1	JC-02-021119-E	217	0.0418	-	7470A	PPB	14 Feb 2019 09:55:14	U	SMPL
K1343-12 - 1	JC-02-021119-F	216	0.0416	-	7470A	PPB	14 Feb 2019 09:57:21	U	SMPL
K1343-14 - 1	JC-02-021119-G	283	0.0583	-	7470A	PPB	14 Feb 2019 09:59:28	U	SMPL
K1343-16 - 1	JC-02-021119-H	271	0.0553	-	7470A	PPB	14 Feb 2019 10:01:35	U	SMPL
K1343-18 - 1	JC-02-021119-I	221	0.0428	-	7470A	PPB	14 Feb 2019 10:03:43	U	SMPL
K1343-20 - 1	JC-02-021119-J	255	0.0513	-	7470A	PPB	14 Feb 2019 10:05:51	U	SMPL
K1343-22 - 1	JC-02-021119-K	219	0.0423	-	7470A	PPB	14 Feb 2019 10:07:58	U	SMPL
K1343-24 - 1	JC-02-021119-L	206	0.0391	-	7470A	PPB	14 Feb 2019 10:10:05	U	SMPL
CCV53 - 1	CCV53	19702	4.8902	-	7470A	PPB	14 Feb 2019 10:12:12	U	SMPL
CCB53 - 1	CCB53	-32	-0.0201	-	7470A	PPB	14 Feb 2019 10:14:18	U	SMPL
K1343-26 - 1	JC-02-021119-M	295	0.0612	-	7470A	PPB	14 Feb 2019 10:16:24	U	SMPL
K1343-28 - 1	JC-02-021119-N	298	0.0620	-	7470A	PPB	14 Feb 2019 10:18:31	U	SMPL
K1343-30 - 1	JC-02-021119-O	184	0.0336	-	7470A	PPB	14 Feb 2019 10:20:38	U	SMPL
K1343-32 - 1	JC-02-021119-P	37	-0.0030	-	7470A	PPB	14 Feb 2019 10:22:44	U	SMPL
PB117098TB - 1	PB117062TB	189	0.0349	-	7470A	PPB	14 Feb 2019 10:24:51	U	SMPL
K1343-02LX5 - 1		49	0.0000	-	7470A	PPB	14 Feb 2019 10:26:58	U	SMPL
K1343-02A - 1		17912	4.4448	-	7470A	PPB	14 Feb 2019 10:29:05	U	SMPL
CCV54 - 1	CCV54	19585	4.8611	-	7470A	PPB	14 Feb 2019 10:31:13	U	SMPL
CCB54 - 1	CCB54	-28	-0.0191	-	7470A	PPB	14 Feb 2019 10:33:21	U	SMPL