


LB100832

7471B

INSTRUMENT ID : CV1

 μ Abs. :

56879



Linear

A= 0.0000e+000

B= 1.7518e-004

C= 5.4796e-002

Rho= 0.9998732

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.057	0.057	14	0.000	14				
0.2	0.200	0.231	0.031	1004	0.0 %	1004				
2.5	2.500	2.427	-0.073	13542	0.0 %	13542				
5.0	5.000	4.912	-0.088	27725	0.0 %	27725				
7.5	7.500	7.555	0.055	42814	0.0 %	42814				
10.0	10.000	10.019	0.019	56879	0.0 %	56879				

LB100832 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 14 Feb 2019 11:21:28

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	14	-	0.0000	7471B	PPB	14 Feb 2019 11:24:15	S	Std
0.2 - 1	50.2	1004	-	0.2000	7471B	PPB	14 Feb 2019 11:26:33	S	Std
2.5 - 1	52.5	13542	-	2.5000	7471B	PPB	14 Feb 2019 11:28:50	S	Std
5.0 - 1	55.0	27725	-	5.0000	7471B	PPB	14 Feb 2019 11:31:07	S	Std
7.5 - 1	57.5	42814	-	7.5000	7471B	PPB	14 Feb 2019 11:33:24	S	Std
10.0 - 1	510.0	56879	-	10.0000	7471B	PPB	14 Feb 2019 11:35:41	S	Std
ICV10 - 1	ICV10	21663	3.8496	-	7471B	PPB	14 Feb 2019 11:45:38	U	SMPL
ICB10 - 1	ICB10	-56	0.0450	-	7471B	PPB	14 Feb 2019 11:47:53	U	SMPL
CCV55 - 1	CCV55	27728	4.9121	-	7471B	PPB	14 Feb 2019 11:50:09	U	SMPL
CCB55 - 1	CCB55	-69	0.0427	-	7471B	PPB	14 Feb 2019 11:52:24	U	SMPL
CRA - 1	CRA	955	0.2221	-	7471B	PPB	14 Feb 2019 11:54:40	U	SMPL
HighStd - 1	HighStd	57107	10.0586	-	7471B	PPB	14 Feb 2019 11:56:55	U	SMPL
ChkStd - 1	ChkStd	39633	6.9976	-	7471B	PPB	14 Feb 2019 11:59:10	U	SMPL
PB117109BL - 1	PBS	0	0.0548	-	7471B	PPB	14 Feb 2019 12:39:01	U	SMPL
PB117109BS - 1	LCSS	19426	3.4578	-	7471B	PPB	14 Feb 2019 12:41:16	U	SMPL
K1349-05 - 1	OR-01-021319-A	7237	1.3225	-	7471B	PPB	14 Feb 2019 12:43:33	U	SMPL
K1349-07 - 1	OR-01-021319-B	1855	0.3797	-	7471B	PPB	14 Feb 2019 12:45:50	U	SMPL
K1350-05 - 1	EO-02-021319-A	1976	0.4009	-	7471B	PPB	14 Feb 2019 12:50:35	U	SMPL
K1350-07 - 1	EO-02-021319-B	3695	0.7021	-	7471B	PPB	14 Feb 2019 12:52:52	U	SMPL
K1350-09 - 1	EO-02-021319-C	2505	0.4936	-	7471B	PPB	14 Feb 2019 12:55:07	U	SMPL
CCV56 - 1	CCV56	27837	4.9312	-	7471B	PPB	14 Feb 2019 12:57:22	U	SMPL
CCB56 - 1	CCB56	-56	0.0450	-	7471B	PPB	14 Feb 2019 12:59:38	U	SMPL
K1350-09DUP - 1	EO-02-021319-CDUP	2535	0.4989	-	7471B	PPB	14 Feb 2019 13:01:54	U	SMPL
K1350-09MS - 1	EO-02-021319-CMS	23201	4.1191	-	7471B	PPB	14 Feb 2019 13:04:09	U	SMPL
K1350-09MSD - 1	EO-02-021319-CMSD	23297	4.1359	-	7471B	PPB	14 Feb 2019 13:06:25	U	SMPL
K1350-09LX5 - 1		441	0.1320	-	7471B	PPB	14 Feb 2019 13:08:42	U	SMPL
K1350-09A - 1		25501	4.5220	-	7471B	PPB	14 Feb 2019 13:22:09	U	SMPL
CCV57 - 1	CCV57	28047	4.9680	-	7471B	PPB	14 Feb 2019 13:24:25	U	SMPL
CCB57 - 1	CCB57	-60	0.0443	-	7471B	PPB	14 Feb 2019 13:26:42	U	SMPL