



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 2/15/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 15:35
In Date: 02/14/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:25
Out Date: 02/15/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB100835

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K1491-01	CB571	1	1.15	9.55	7.25	72.6
K1491-02	CB572	2	1.15	9.95	7.73	74.8
K1491-03	CB573	3	1.15	9.59	6.22	60.1
K1491-04	CB574	4	1.15	9.84	4.09	33.8
K1491-05	CB575	5	1.13	9.71	6.95	67.8
K1491-06	CB576	6	1.14	9.69	7.02	68.8
K1491-07	CB577	7	1.17	9.82	7.45	72.6
K1491-08	CB578	8	1.15	9.95	8.16	79.7
K1491-09	CB579	9	1.17	9.96	7.56	72.7
K1491-10	CB580	10	1.16	9.94	4.59	39.1
K1491-11	CB581	11	1.12	9.73	6.66	64.3
K1491-12	CB581MS	12	1.12	9.73	6.66	64.3
K1491-13	CB581MSD	13	1.12	9.73	6.66	64.3
K1491-14	CB582	14	1.18	9.97	6.63	62.0
K1491-15	CB583	15	1.15	9.93	6.41	59.9
K1491-16	CB584	16	1.19	9.83	7.03	67.6
K1491-17	CB5A3	17	1.13	9.94	7.95	77.4
K1491-18	CB5A4	18	1.15	9.98	7.44	71.2
K1491-19	CB5A5	19	1.17	9.73	6.69	64.5
K1491-20	CB5A6	20	1.14	9.78	7.65	75.3
K1491-21	CB5A7	21	1.11	9.62	6.65	65.1
K1491-22	CB5A8	22	1.17	9.72	7.29	71.6

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$