

LB100862


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

40773



A= 0.0000e+000

B= 2.4383e-004

C= -1.6937e-003

Rho= 0.9999227

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.001	-0.001	1	0.000	1				
0.2	0.200	0.178	-0.022	739	0.0 %	739				
2.5	2.500	2.517	0.017	10330	0.0 %	10330				
5.0	5.000	4.978	-0.022	20423	0.0 %	20423				
7.5	7.500	7.588	0.088	31128	0.0 %	31128				
10.0	10.000	9.940	-0.060	40773	0.0 %	40773				

LB100862 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Feb 2019 13:59:01

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	1	-	0.00007470A	PPB		15 Feb 2019 14:12:08	S	Std
0.2 - 1	50.2	739	-	0.20007470A	PPB		15 Feb 2019 14:14:16	S	Std
2.5 - 1	52.5	10330	-	2.50007470A	PPB		15 Feb 2019 14:16:25	S	Std
5.0 - 1	55.0	20423	-	5.00007470A	PPB		15 Feb 2019 14:18:33	S	Std
7.5 - 1	57.5	31128	-	7.50007470A	PPB		15 Feb 2019 14:26:40	S	Std
10.0 - 1	510.0	40773	-	10.00007470A	PPB		15 Feb 2019 14:31:13	S	Std
ICV12 - 1	ICV12	15962	3.8903	-7470A	PPB		15 Feb 2019 14:34:15	U	SMPL
ICB12 - 1	ICB12	-34	-0.0100	-7470A	PPB		15 Feb 2019 14:36:21	U	SMPL
CCV62 - 1	CCV62	20419	4.9770	-7470A	PPB		15 Feb 2019 14:38:27	U	SMPL
CCB62 - 1	CCB62	-27	-0.0083	-7470A	PPB		15 Feb 2019 14:40:33	U	SMPL
CRA - 1	CRA	806	0.1948	-7470A	PPB		15 Feb 2019 14:48:20	U	SMPL
HighStd - 1	HighStd	40859	9.9608	-7470A	PPB		15 Feb 2019 14:50:25	U	SMPL
ChkStd - 1	ChkStd	28122	6.8552	-7470A	PPB		15 Feb 2019 14:52:32	U	SMPL
PB117117BL - 1	PBW	-56	-0.0153	-7470A	PPB		15 Feb 2019 14:54:40	U	SMPL
PB117117BS - 1	LCSW	15645	3.8130	-7470A	PPB		15 Feb 2019 14:56:47	U	SMPL
K1349-06 - 1	OR-01-021319-A	-9	-0.0039	-7470A	PPB		15 Feb 2019 14:58:55	U	SMPL
K1349-06DUP - 1	OR-01-021319-ADUP	28	0.0051	-7470A	PPB		15 Feb 2019 15:01:02	U	SMPL
K1349-06MS - 1	OR-01-021319-AMS	15605	3.8032	-7470A	PPB		15 Feb 2019 15:03:10	U	SMPL
K1349-06MSD - 1	OR-01-021319-AMSD	15817	3.8549	-7470A	PPB		15 Feb 2019 15:05:18	U	SMPL
K1349-08 - 1	OR-01-021319-B	-4	-0.0027	-7470A	PPB		15 Feb 2019 15:07:24	U	SMPL
CCV63 - 1	CCV63	19974	4.8685	-7470A	PPB		15 Feb 2019 15:09:30	U	SMPL
CCB63 - 1	CCB63	-30	-0.0090	-7470A	PPB		15 Feb 2019 15:11:36	U	SMPL
K1350-06 - 1	EO-02-021319-A	21	0.0034	-7470A	PPB		15 Feb 2019 15:13:43	U	SMPL
K1350-08 - 1	EO-02-021319-B	12	0.0012	-7470A	PPB		15 Feb 2019 15:15:49	U	SMPL
K1350-10 - 1	EO-02-021319-C	44	0.0090	-7470A	PPB		15 Feb 2019 15:17:55	U	SMPL
K1479-01 - 1	OIL-WATER-SEPERATOR	31	0.0059	-7470A	PPB		15 Feb 2019 15:20:02	U	SMPL
K1480-01 - 1	OIL-WATER-SEPERATOR-3	20	0.0032	-7470A	PPB		15 Feb 2019 15:22:09	U	SMPL
PB117118BL - 1	PBW	6	-0.0002	-7470A	PPB		15 Feb 2019 15:24:16	U	SMPL
PB117118BS - 1	LCSW	16258	3.9624	-7470A	PPB		15 Feb 2019 15:26:24	U	SMPL
K1458-03 - 1	GW-1	-5	-0.0029	-7470A	PPB		15 Feb 2019 15:28:32	U	SMPL
K1458-03DUP - 1	GW-1DUP	28	0.0051	-7470A	PPB		15 Feb 2019 15:30:39	U	SMPL
K1458-03MS - 1	GW-1MS	15294	3.7274	-7470A	PPB		15 Feb 2019 15:32:45	U	SMPL
CCV64 - 1	CCV64	20242	4.9338	-7470A	PPB		15 Feb 2019 15:34:52	U	SMPL
CCB64 - 1	CCB64	-37	-0.0107	-7470A	PPB		15 Feb 2019 15:36:58	U	SMPL
K1458-03MSD - 1	GW-1MSD	15556	3.7913	-7470A	PPB		15 Feb 2019 15:39:05	U	SMPL
K1458-12 - 1	GW-7	-20	-0.0066	-7470A	PPB		15 Feb 2019 15:41:11	U	SMPL
K1458-17 - 1	GW-1	25	0.0044	-7470A	PPB		15 Feb 2019 15:43:17	U	SMPL
K1458-18 - 1	GW-7	48	0.0100	-7470A	PPB		15 Feb 2019 15:45:24	U	SMPL
PB117151BL - 1	PBW	181	0.0424	-7470A	PPB		15 Feb 2019 15:47:30	U	SMPL
PB117151BS - 1	LCSW	15439	3.7627	-7470A	PPB		15 Feb 2019 15:49:38	U	SMPL
K1346-08 - 1	CL-01-021419-A	179	0.0420	-7470A	PPB		15 Feb 2019 15:51:45	U	SMPL
K1346-08DUP - 1	CL-01-021419-ADUP	192	0.0451	-7470A	PPB		15 Feb 2019 15:53:53	U	SMPL
K1346-08MS - 1	CL-01-021419-AMS	14562	3.5489	-7470A	PPB		15 Feb 2019 15:56:01	U	SMPL
K1346-08MSD - 1	CL-01-021419-AMSD	15704	3.8274	-7470A	PPB		15 Feb 2019 15:58:08	U	SMPL
CCV65 - 1	CCV65	21368	5.2084	-7470A	PPB		15 Feb 2019 16:00:14	U	SMPL
CCB65 - 1	CCB65	-40	-0.0114	-7470A	PPB		15 Feb 2019 16:02:20	U	SMPL
K1346-10 - 1	CL-01-021419-B	206	0.0485	-7470A	PPB		15 Feb 2019 16:04:26	U	SMPL
K1346-12 - 1	CL-01-021419-C	209	0.0493	-7470A	PPB		15 Feb 2019 16:06:33	U	SMPL
K1346-14 - 1	CL-01-021419-D	217	0.0512	-7470A	PPB		15 Feb 2019 16:08:40	U	SMPL
K1346-16 - 1	CL-01-021419-E	219	0.0517	-7470A	PPB		15 Feb 2019 16:10:46	U	SMPL
K1358-04 - 1	TR-04-021419	201	0.0473	-7470A	PPB		15 Feb 2019 16:12:54	U	SMPL
K1484-04 - 1	M-2-S	30	0.0056	-7470A	PPB		15 Feb 2019 16:15:01	U	SMPL
K1484-08 - 1	M-2-D	160	0.0373	-7470A	PPB		15 Feb 2019 16:17:08	U	SMPL
K1484-12 - 1	M-3-S	224	0.0529	-7470A	PPB		15 Feb 2019 16:19:16	U	SMPL
K1484-16 - 1	M-3-D	243	0.0576	-7470A	PPB		15 Feb 2019 16:21:24	U	SMPL
K1486-03 - 1	VNJ-227	210	0.0495	-7470A	PPB		15 Feb 2019 16:23:31	U	SMPL
CCV66 - 1	CCV66	21069	5.1355	-7470A	PPB		15 Feb 2019 16:25:37	U	SMPL
CCB66 - 1	CCB66	-25	-0.0078	-7470A	PPB		15 Feb 2019 16:27:44	U	SMPL
K1490-04 - 1	TP-1-214	198	0.0466	-7470A	PPB		15 Feb 2019 16:29:50	U	SMPL
K1500-03 - 1	G-021419	176	0.0412	-7470A	PPB		15 Feb 2019 16:31:57	U	SMPL
K1501-03 - 1	CH-021419	182	0.0427	-7470A	PPB		15 Feb 2019 16:34:03	U	SMPL
K1502-03 - 1	RO-1-4	229	0.0541	-7470A	PPB		15 Feb 2019 16:36:10	U	SMPL
PB117152BL - 1	PBW	191	0.0449	-7470A	PPB		15 Feb 2019 16:38:16	U	SMPL
PB117152BS - 1	LCSW	15917	3.8793	-7470A	PPB		15 Feb 2019 16:40:24	U	SMPL
K1502-04 - 1	DRUM-A-1-5	164	0.0383	-7470A	PPB		15 Feb 2019 16:42:31	U	SMPL
K1502-05 - 1	DRUM-B-1-5	181	0.0424	-7470A	PPB		15 Feb 2019 16:44:39	U	SMPL
K1502-06 - 1	DRUM-C-1-5	197	0.0463	-7470A	PPB		15 Feb 2019 16:46:47	U	SMPL
K1502-07 - 1	DRUM-D-1-5	196	0.0461	-7470A	PPB		15 Feb 2019 16:48:54	U	SMPL
CCV67 - 1	CCV67	20828	5.0767	-7470A	PPB		15 Feb 2019 16:51:01	U	SMPL
CCB67 - 1	CCB67	-21	-0.0068	-7470A	PPB		15 Feb 2019 16:53:08	U	SMPL
K1502-08 - 1	DRUM-E-1-5	175	0.0410	-7470A	PPB		15 Feb 2019 16:55:15	U	SMPL
K1502-09 - 1	DRUM-F-1-5	183	0.0429	-7470A	PPB		15 Feb 2019 16:57:22	U	SMPL
K1502-10 - 1	DRUM-G-1-5	189	0.0444	-7470A	PPB		15 Feb 2019 16:59:30	U	SMPL
K1502-11 - 1	DRUM-H-1-5	171	0.0400	-7470A	PPB		15 Feb 2019 17:01:37	U	SMPL
K1502-12 - 1	DRUM-I-1-5	197	0.0463	-7470A	PPB		15 Feb 2019 17:03:44	U	SMPL
K1502-13 - 1	DRUM-J-1-5	202	0.0476	-7470A	PPB		15 Feb 2019 17:05:51	U	SMPL
K1502-13DUP - 1	DRUM-J-1-5DUP	192	0.0451	-7470A	PPB		15 Feb 2019 17:07:59	U	SMPL
K1502-13MS - 1	DRUM-J-1-5MS	16078	3.9185	-7470A	PPB		15 Feb 2019 17:10:07	U	SMPL
K1502-13MSD - 1	DRUM-J-1-5MSD	16072	3.9171	-7470A	PPB		15 Feb 2019 17:12:15	U	SMPL
PB117117TB - 1	PB117096TB	22	0.0037	-7470A	PPB		15 Feb 2019 17:14:22	U	SMPL
CCV68 - 1	CCV68	20526	5.0031	-7470A	PPB		15 Feb 2019 17:16:29	U	SMPL

LB100862 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Feb 2019 13:59:01

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB68 - 1	CCB68	102	0.0232		-7470A	PPB	15 Feb 2019 17:18:36	U	SMPL
PB117151TB - 1	PB117123TB	193	0.0454		-7470A	PPB	15 Feb 2019 17:20:43	U	SMPL
PB117152TB - 1	PB117144TB	55	0.0117		-7470A	PPB	15 Feb 2019 17:22:50	U	SMPL
K1349-06LX5 - 1		111	0.0254		-7470A	PPB	15 Feb 2019 17:24:57	U	SMPL
K1349-06A - 1		16873	4.1124		-7470A	PPB	15 Feb 2019 17:27:04	U	SMPL
K1458-03LX5 - 1		99	0.0224		-7470A	PPB	15 Feb 2019 17:29:12	U	SMPL
K1458-03A - 1		16910	4.1214		-7470A	PPB	15 Feb 2019 17:31:19	U	SMPL
K1346-08LX5 - 1		121	0.0278		-7470A	PPB	15 Feb 2019 17:33:26	U	SMPL
K1346-08A - 1		14960	3.6459		-7470A	PPB	15 Feb 2019 17:35:34	U	SMPL
K1502-13LX5 - 1		128	0.0295		-7470A	PPB	15 Feb 2019 17:37:42	U	SMPL
K1502-13A - 1		16348	3.9844		-7470A	PPB	15 Feb 2019 17:39:49	U	SMPL
CCV69 - 1	CCV69	21033	5.1267		-7470A	PPB	15 Feb 2019 17:41:56	U	SMPL
CCB69 - 1	CCB69	-39	-0.0112		-7470A	PPB	15 Feb 2019 17:44:03	U	SMPL