



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB100881

Analyst By : JIGNESH

Supervisor Review By : apatel

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	02/15/2019	16:30
2	CAL2	1	Water	NA	NA	20.1	7.01	02/15/2019	16:31
3	CAL3	1	Water	NA	NA	20.1	10.02	02/15/2019	16:32
4	ICV	1	Water	NA	NA	20.2	6.99	02/15/2019	16:34
5	CCV1	1	Water	NA	NA	20.2	2.02	02/15/2019	16:35
6	K1519-01	1	Water	NA	NA	16.2	7.51	02/15/2019	16:36
7	K1519-01DUP	1	Water	NA	NA	16.3	7.51	02/15/2019	16:37
8	CCV2	1	Water	NA	NA	20.1	12.01	02/15/2019	16:38

23 100881

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY WATER I

WorkList ID : 122433

Date : 02-15-2019 16:20:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/15/2019	Water	K1519-01	Corrosivity	Cool 4 deg C	LANG01	M51	005-WL-1	02/15/2019	9040C

Date/Time 02-15-19 16:25
 Received by: Jo
 Relinquished by: [Signature]

Date/Time 02-15-19 17:00
 Received by: [Signature]
 Relinquished by: [Signature]