



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB100885

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.1	4.01	02/18/2019	15:20
2	CAL2	1	Water	NA	NA	22.1	7.01	02/18/2019	15:21
3	CAL3	1	Water	NA	NA	22.1	10.02	02/18/2019	15:22
4	ICV	1	Water	NA	NA	22.1	6.99	02/18/2019	15:24
5	CCV1	1	Water	NA	NA	22.1	2.01	02/18/2019	15:25
6	K1525-04	1	Solid	20.02	20	21.8	7.75	02/18/2019	15:26
7	K1525-04DUP	1	Solid	20.03	20	21.9	7.76	02/18/2019	15:27
8	K1525-08	1	Solid	20.04	20	21.7	7.14	02/18/2019	15:28
9	CCV2	1	Water	NA	NA	22.2	12.01	02/18/2019	15:30

2B100885

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K1525

WorkList ID : 122442

Date : 02-18-2019 11:59:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/15/2019	Solid	K1525-04	Corrosivity	Cool 4 deg C	PSEG01	M51	M-14-S	02/15/2019	9045D
02/15/2019	Solid	K1525-08	Corrosivity	Cool 4 deg C	PSEG01	M51	M-14-D	02/15/2019	9045D

Date/Time 02-18-19 15:15
Received by: SE
Relinquished by: CS

Date/Time 02-18-19 16:25
Received by: SE
Relinquished by: CS