

LB100893


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

47800



A= 0.0000e+000

B= 2.0887e-004

C= -2.2168e-002

Rho= 0.9999232

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.012	-0.012	50	0.000	50				
0.2	0.200	0.237	0.037	1243	0.0 %	1243				
2.5	2.500	2.464	-0.036	11904	0.0 %	11904				
5.0	5.000	4.964	-0.036	23873	0.0 %	23873				
7.5	7.500	7.584	0.084	36414	0.0 %	36414				
10.0	10.000	9.962	-0.038	47800	0.0 %	47800				

LB100893 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Feb 2019 09:58:35

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	50	-	0.0007470A	PPB		19 Feb 2019 09:59:24	S	Std
0.2 - 1	50.2	1243	-	0.20007470A	PPB		19 Feb 2019 10:01:32	S	Std
2.5 - 1	52.5	11904	-	2.50007470A	PPB		19 Feb 2019 10:03:40	S	Std
5.0 - 1	55.0	23873	-	5.00007470A	PPB		19 Feb 2019 10:05:48	S	Std
7.5 - 1	57.5	36414	-	7.50007470A	PPB		19 Feb 2019 10:07:55	S	Std
10.0 - 1	510.0	47800	-	10.00007470A	PPB		19 Feb 2019 10:10:02	S	Std
ICV14 - 1	ICV14	19541	4.0594	-7470A	PPB		19 Feb 2019 10:16:34	U	SMPL
ICB14 - 1	ICB14	5	-0.0211	-7470A	PPB		19 Feb 2019 10:18:41	U	SMPL
CCV78 - 1	CCV78	23889	4.9676	-7470A	PPB		19 Feb 2019 10:20:47	U	SMPL
CCB78 - 1	CCB78	-6	-0.0234	-7470A	PPB		19 Feb 2019 10:22:52	U	SMPL
CRA - 1	CRA	1202	0.2289	-7470A	PPB		19 Feb 2019 10:31:19	U	SMPL
HighStd - 1	HighStd	47326	9.8630	-7470A	PPB		19 Feb 2019 10:33:25	U	SMPL
ChkStd - 1	ChkStd	33356	6.9450	-7470A	PPB		19 Feb 2019 10:35:31	U	SMPL
PB117203BL - 1	PBW	-63	-0.0353	-7470A	PPB		19 Feb 2019 10:37:38	U	SMPL
PB117203BS - 1	LCSW	17764	3.6883	-7470A	PPB		19 Feb 2019 10:42:26	U	SMPL
K1519-01 - 1	005-WL-1	-6	-0.0234	-7470A	PPB		19 Feb 2019 10:44:33	U	SMPL
K1519-01DUP - 1	005-WL-1DUP	35	-0.0149	-7470A	PPB		19 Feb 2019 10:46:40	U	SMPL
K1519-01MS - 1	005-WL-1MS	16814	3.4898	-7470A	PPB		19 Feb 2019 10:48:48	U	SMPL
K1519-01MSD - 1	005-WL-1MSD	16851	3.4976	-7470A	PPB		19 Feb 2019 10:50:56	U	SMPL
PB117204BL - 1	PBW	-28	-0.0280	-7470A	PPB		19 Feb 2019 10:53:02	U	SMPL
CCV79 - 1	CCV79	23843	4.9580	-7470A	PPB		19 Feb 2019 10:55:09	U	SMPL
CCB79 - 1	CCB79	-31	-0.0286	-7470A	PPB		19 Feb 2019 10:57:14	U	SMPL
PB117204BS - 1	LCSW	16565	3.4378	-7470A	PPB		19 Feb 2019 10:59:21	U	SMPL
K1508-04 - 1	M-12-S	-11	-0.0245	-7470A	PPB		19 Feb 2019 11:01:27	U	SMPL
K1508-04DUP - 1	M-12-SDUP	21	-0.0178	-7470A	PPB		19 Feb 2019 11:03:33	U	SMPL
K1508-04MS - 1	M-12-SMS	16772	3.4811	-7470A	PPB		19 Feb 2019 11:05:39	U	SMPL
K1508-04MSD - 1	M-12-SMSD	16739	3.4742	-7470A	PPB		19 Feb 2019 11:07:46	U	SMPL
K1508-08 - 1	M-12-D	-43	-0.0311	-7470A	PPB		19 Feb 2019 11:09:53	U	SMPL
K1508-12 - 1	M-13-S	10	-0.0201	-7470A	PPB		19 Feb 2019 11:12:00	U	SMPL
K1508-16 - 1	M-13-D	13	-0.0195	-7470A	PPB		19 Feb 2019 11:14:08	U	SMPL
K1511-02 - 1	STONES	16	-0.0188	-7470A	PPB		19 Feb 2019 11:16:17	U	SMPL
K1513-02 - 1	FLOORS	403	0.0620	-7470A	PPB		19 Feb 2019 11:18:23	U	SMPL
CCV80 - 1	CCV80	23827	4.9547	-7470A	PPB		19 Feb 2019 11:20:29	U	SMPL
CCB80 - 1	CCB80	-29	-0.0282	-7470A	PPB		19 Feb 2019 11:22:36	U	SMPL
K1513-04 - 1	WALL-ONE	38	-0.0142	-7470A	PPB		19 Feb 2019 11:24:42	U	SMPL
K1513-06 - 1	INNER-WALL	55	-0.0107	-7470A	PPB		19 Feb 2019 11:26:49	U	SMPL
K1517-10 - 1	G-021519	29	-0.0161	-7470A	PPB		19 Feb 2019 11:28:55	U	SMPL
K1518-02 - 1	CABLE-BRIDGE-FOUNDATION-D-1	26	-0.0167	-7470A	PPB		19 Feb 2019 11:31:02	U	SMPL
K1518-04 - 1	CABLE-BRIDGE-FOUNDATION-D-2	8	-0.0205	-7470A	PPB		19 Feb 2019 11:33:09	U	SMPL
K1518-06 - 1	CABLE-BRIDGE-FOUNDATION-D-3	39	-0.0140	-7470A	PPB		19 Feb 2019 11:35:16	U	SMPL
K1518-08 - 1	CABLE-BRIDGE-FOUNDATION-D-4	40	-0.0138	-7470A	PPB		19 Feb 2019 11:37:24	U	SMPL
K1518-10 - 1	CABLE-BRIDGE-FOUNDATION-D-5	30	-0.0159	-7470A	PPB		19 Feb 2019 11:39:32	U	SMPL
K1518-12 - 1	CABLE-BRIDGE-FOUNDATION-D-6	54	-0.0109	-7470A	PPB		19 Feb 2019 11:41:39	U	SMPL
K1518-14 - 1	CABLE-BRIDGE-FOUNDATION-D-7	18	-0.0184	-7470A	PPB		19 Feb 2019 11:43:46	U	SMPL
CCV81 - 1	CCV81	23834	4.9561	-7470A	PPB		19 Feb 2019 11:45:52	U	SMPL
CCB81 - 1	CCB81	-25	-0.0274	-7470A	PPB		19 Feb 2019 11:47:59	U	SMPL
K1518-16 - 1	CABLE-BRIDGE-FOUNDATION-D-8	8	-0.0205	-7470A	PPB		19 Feb 2019 11:50:05	U	SMPL
K1518-18 - 1	TP-MH-C02	26	-0.0167	-7470A	PPB		19 Feb 2019 11:52:12	U	SMPL
K1518-20 - 1	TP-MH-C05	16	-0.0188	-7470A	PPB		19 Feb 2019 11:54:18	U	SMPL
K1518-22 - 1	CABLE-BRIDGE-FOUNDATION-B-10	22	-0.0176	-7470A	PPB		19 Feb 2019 11:56:25	U	SMPL
PB117205BL - 1	PBW	115	0.0019	-7470A	PPB		19 Feb 2019 11:58:32	U	SMPL
PB117205BS - 1	LCSW	17343	3.6003	-7470A	PPB		19 Feb 2019 12:00:39	U	SMPL
K1512-01 - 1	TankA_190214	171	0.0135	-7470A	PPB		19 Feb 2019 12:02:46	U	SMPL
K1512-01DUP - 1	TankA_190214DUP	201	0.0198	-7470A	PPB		19 Feb 2019 12:04:54	U	SMPL
K1512-01MS - 1	TankA_190214MS	16659	3.4575	-7470A	PPB		19 Feb 2019 12:07:02	U	SMPL
K1512-01MSD - 1	TankA_190214MSD	16353	3.3935	-7470A	PPB		19 Feb 2019 12:09:08	U	SMPL
CCV82 - 1	CCV82	23001	4.7821	-7470A	PPB		19 Feb 2019 12:11:15	U	SMPL
CCB82 - 1	CCB82	-45	-0.0316	-7470A	PPB		19 Feb 2019 12:13:21	U	SMPL
K1512-02 - 1	TankB_190214	147	0.0085	-7470A	PPB		19 Feb 2019 12:15:27	U	SMPL
K1512-03 - 1	TankC_190214	168	0.0129	-7470A	PPB		19 Feb 2019 12:17:34	U	SMPL
K1512-04 - 1	TankA_Bottom_190214	16943	3.5168	-7470A	PPB		19 Feb 2019 12:19:40	U	SMPL
K1512-05 - 1	TankB_Bottom_190214	17342	3.6001	-7470A	PPB		19 Feb 2019 12:21:47	U	SMPL
K1512-06 - 1	TankC_Bottom_190214	11298	2.3377	-7470A	PPB		19 Feb 2019 12:23:54	U	SMPL
K1512-07 - 1	TankA_190214	235	0.0269	-7470A	PPB		19 Feb 2019 12:26:00	U	SMPL
K1512-08 - 1	TankB_190214	243	0.0286	-7470A	PPB		19 Feb 2019 12:28:08	U	SMPL
K1512-09 - 1	TankC_190214	240	0.0280	-7470A	PPB		19 Feb 2019 12:30:15	U	SMPL
K1512-10 - 1	TankA_Bottom_190214	281	0.0365	-7470A	PPB		19 Feb 2019 12:32:23	U	SMPL
K1512-11 - 1	TankB_Bottom_190214	95	-0.0023	-7470A	PPB		19 Feb 2019 12:34:29	U	SMPL
CCV83 - 1	CCV83	23260	4.8362	-7470A	PPB		19 Feb 2019 12:36:37	U	SMPL
CCB83 - 1	CCB83	-50	-0.0326	-7470A	PPB		19 Feb 2019 12:38:43	U	SMPL
K1512-12 - 1	TankC_Bottom_190214	28	-0.0163	-7470A	PPB		19 Feb 2019 12:40:50	U	SMPL
PB117203TB - 1	PB117185TB	6	-0.0209	-7470A	PPB		19 Feb 2019 12:42:58	U	SMPL
PB117204TB - 1	PB117158TB	35	-0.0149	-7470A	PPB		19 Feb 2019 12:45:05	U	SMPL
K1519-01LX5 - 1		-2	-0.0226	-7470A	PPB		19 Feb 2019 12:47:12	U	SMPL
K1519-01A - 1		17092	3.5479	-7470A	PPB		19 Feb 2019 12:49:19	U	SMPL
K1508-04LX5 - 1		-23	-0.0270	-7470A	PPB		19 Feb 2019 12:51:27	U	SMPL
K1508-04A - 1		16784	3.4836	-7470A	PPB		19 Feb 2019 12:53:34	U	SMPL
K1512-01LX5 - 1		22	-0.0176	-7470A	PPB		19 Feb 2019 12:55:42	U	SMPL
K1512-01A - 1		16820	3.4911	-7470A	PPB		19 Feb 2019 12:57:50	U	SMPL
CCV84 - 1	CCV84	23324	4.8496	-7470A	PPB		19 Feb 2019 12:59:57	U	SMPL
CCB84 - 1	CCB84	-37	-0.0299	-7470A	PPB		19 Feb 2019 13:02:04	U	SMPL