

LB100896


ISM02.4

INSTRUMENT ID: CV1

Linear

 μ Abs.:

48150



A= 0.0000e+000

B= 2.0817e-004

C= -1.2621e-002

Rho= 0.999378

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.005	-0.005	37	0.000	37					
0.2	0.200	0.245	0.045	1236	0.0 %	1236					
2.5	2.500	2.489	-0.011	12019	0.0 %	12019					
5.0	5.000	4.920	-0.080	23697	0.0 %	23697					
7.5	7.500	7.540	0.040	36280	0.0 %	36280					
10.0	10.000	10.011	0.011	48150	0.0 %	48150					

0-20122

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.00 - 1		37	-		0 ISM02.4	PPB	2/19/2019 14:04	S	Std
0.2 - 1		1236	-		0.2 ISM02.4	PPB	2/19/2019 14:06	S	Std
2.5 - 1		12019	-		2.5 ISM02.4	PPB	2/19/2019 14:09	S	Std
5.0 - 1		23697	-		5 ISM02.4	PPB	2/19/2019 14:11	S	Std
7.5 - 1		36280	-		7.5 ISM02.4	PPB	2/19/2019 14:21	S	Std
10.0 - 1		48150	-		10 ISM02.4	PPB	2/19/2019 14:23	S	Std
ICV - 1	ICV	18036	3.7419 -		ISM02.4	PPB	2/19/2019 14:26	U	SMPL
ICB - 1	ICB	22	-0.008 -		ISM02.4	PPB	2/19/2019 14:28	U	SMPL
CCV085 - 1	CCV085	23923	4.9674 -		ISM02.4	PPB	2/19/2019 14:31	U	SMPL
CCB085 - 1	CCB085	92	0.0065 -		ISM02.4	PPB	2/19/2019 14:33	U	SMPL
PB117223BL - 1	PBW223	148	0.0182 -		ISM02.4	PPB	2/19/2019 14:38	U	SMPL
K1439-01 - 1	MESXA4	183	0.0255 -		ISM02.4	PPB	2/19/2019 14:40	U	SMPL
K1439-02 - 1	MESXA7	170	0.0228 -		ISM02.4	PPB	2/19/2019 14:42	U	SMPL
K1439-03 - 1	MESXD4	131	0.0146 -		ISM02.4	PPB	2/19/2019 14:45	U	SMPL
K1439-04 - 1	MESXD6	164	0.0215 -		ISM02.4	PPB	2/19/2019 14:47	U	SMPL
K1439-05 - 1	MESXD6D	200	0.029 -		ISM02.4	PPB	2/19/2019 14:49	U	SMPL
K1439-06 - 1	MESXD6S	4953	1.0184 -		ISM02.4	PPB	2/19/2019 14:52	U	SMPL
K1439-07 - 1	MESXF7	152	0.019 -		ISM02.4	PPB	2/19/2019 14:54	U	SMPL
PB117224BL - 1	PBW224	156	0.0199 -		ISM02.4	PPB	2/19/2019 14:56	U	SMPL
K1521-01 - 1	MCOAM8	397	0.07 -		ISM02.4	PPB	2/19/2019 14:58	U	SMPL
K1521-02 - 1	MCOAM8D	370	0.0644 -		ISM02.4	PPB	2/19/2019 15:04	U	SMPL
K1521-03 - 1	MCOAM8S	4769	0.9801 -		ISM02.4	PPB	2/19/2019 15:07	U	SMPL
K1521-04 - 1	MCOAM9	197	0.0284 -		ISM02.4	PPB	2/19/2019 15:09	U	SMPL
CCV086 - 1	CCV086	23596	4.8993 -		ISM02.4	PPB	2/19/2019 15:11	U	SMPL
CCB086 - 1	CCB086	17	-0.0091 -		ISM02.4	PPB	2/19/2019 15:13	U	SMPL