



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB100904

Ambient Barometric Pressure (mmHg): 750

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	81.2	02/19/2019	16:25
2	K1519-01		1	8	>100.0	>212.0	>212.0	02/19/2019	16:50
3	K1519-01DUP		1	8	>100.0	>212.0	>212.0	02/19/2019	17:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB100904

Worklist Name : FP21919

Worklist ID : 122471

Date : 02-19-2019 15:58:32

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/20/2019	Water	K1519-01	Flash Point	Cool 4 deg C	LANG01	M51	005-WL-1	02/15/2019	1010A

Date/Time 02/19/19 16:00

Received by: KS

Relinquished by: CP

Date/Time 02/19/19 17:20

Received by: CP

Relinquished by: KS