

LB100910


INSTRUMENT ID: CV1

7470A

Linear

 μ Abs.:

40303



A= 0.0000e+000

B= 2.4633e-004

C= -8.6314e-002

Rho= 0.9995444

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.079	-0.079	30	0.000	30				
0.2	0.200	0.179	-0.021	1077	0.0 %	1077				
2.5	2.500	2.489	-0.011	10456	0.0 %	10456				
5.0	5.000	5.191	0.191	21423	0.0 %	21423				
7.5	7.500	7.578	0.078	31115	0.0 %	31115				
10.0	10.000	9.842	-0.158	40303	0.0 %	40303				

LB100910 INSTRUMENT ID :CV1

Method: 7470A

Operator: Admin

Date of Analysis: 20 Feb 2019 09:26:03

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	30	-	0.0000	7470A	PPB	20 Feb 2019 09:28:39	S	Std
0.2 - 1	50-2	1077	-	0.2000	7470A	PPB	20 Feb 2019 09:30:48	S	Std
2.5 - 1	52-5	10456	-	2.5000	7470A	PPB	20 Feb 2019 09:32:56	S	Std
5.0 - 1	55-0	21423	-	5.0000	7470A	PPB	20 Feb 2019 09:35:04	S	Std
7.5 - 1	57-5	31115	-	7.5000	7470A	PPB	20 Feb 2019 09:44:30	S	Std
10.0 - 1	510-0	40303	-	10.0000	7470A	PPB	20 Feb 2019 09:46:37	S	Std
ICV17 - 1	ICV17	15848	3.8175	-	7470A	PPB	20 Feb 2019 09:49:51	U	SMPL
ICB17 - 1	ICB17	192	-0.0390	-	7470A	PPB	20 Feb 2019 09:51:58	U	SMPL
CCV87 - 1	CCV87	21132	5.1191	-	7470A	PPB	20 Feb 2019 09:54:04	U	SMPL
CCB87 - 1	CCB87	-17	-0.0905	-	7470A	PPB	20 Feb 2019 09:56:10	U	SMPL
CRA - 1	CRA	1076	0.1787	-	7470A	PPB	20 Feb 2019 09:58:16	U	SMPL
HighStd - 1	HighStd	40342	9.8511	-	7470A	PPB	20 Feb 2019 10:00:22	U	SMPL
ChkStd - 1	ChkStd	29356	7.1450	-	7470A	PPB	20 Feb 2019 10:02:28	U	SMPL
PB117240BL - 1	PBW	229	-0.0299	-	7470A	PPB	20 Feb 2019 10:04:36	U	SMPL
PB117240BS - 1	LCSW	15451	3.7197	-	7470A	PPB	20 Feb 2019 10:06:44	U	SMPL
K1528-01 - 1	MW-1	462	0.0275	-	7470A	PPB	20 Feb 2019 10:08:51	U	SMPL
K1528-01DUP - 1	MW-1DUP	435	0.0208	-	7470A	PPB	20 Feb 2019 10:10:58	U	SMPL
K1528-02 - 1	MW-2	714	0.0896	-	7470A	PPB	20 Feb 2019 10:13:06	U	SMPL
K1528-03 - 1	MW-3	466	0.0285	-	7470A	PPB	20 Feb 2019 10:15:14	U	SMPL
K1528-04 - 1	MW-4	543	0.0474	-	7470A	PPB	20 Feb 2019 10:17:20	U	SMPL
CCV88 - 1	CCV88	21106	5.1127	-	7470A	PPB	20 Feb 2019 10:19:26	U	SMPL
CCB88 - 1	CCB88	39	-0.0767	-	7470A	PPB	20 Feb 2019 10:21:33	U	SMPL
K1528-05 - 1	k1528-01MS	18057	4.3617	-	7470A	PPB	20 Feb 2019 10:23:39	U	SMPL
K1528-06 - 1	k1528-01MSD	17860	4.3131	-	7470A	PPB	20 Feb 2019 10:25:45	U	SMPL
K1528-07 - 1	DUPE-X	462	0.0275	-	7470A	PPB	20 Feb 2019 10:27:51	U	SMPL
PB117246BL - 1	PBW	206	-0.0356	-	7470A	PPB	20 Feb 2019 10:29:58	U	SMPL
PB117246BS - 1	LCSW	16753	4.0405	-	7470A	PPB	20 Feb 2019 10:32:04	U	SMPL
K1525-04 - 1	M-14-S	211	-0.0343	-	7470A	PPB	20 Feb 2019 10:34:12	U	SMPL
K1525-04DUP - 1	M-14-SDUP	243	-0.0265	-	7470A	PPB	20 Feb 2019 10:36:19	U	SMPL
K1525-04MS - 1	M-14-SMS	15461	3.7222	-	7470A	PPB	20 Feb 2019 10:38:27	U	SMPL
K1525-04MSD - 1	M-14-SMSD	15291	3.6803	-	7470A	PPB	20 Feb 2019 10:40:35	U	SMPL
K1525-08 - 1	M-14-D	291	-0.0146	-	7470A	PPB	20 Feb 2019 10:42:41	U	SMPL
CCV89 - 1	CCV89	21409	5.1874	-	7470A	PPB	20 Feb 2019 10:44:47	U	SMPL
CCB89 - 1	CCB89	-24	-0.0922	-	7470A	PPB	20 Feb 2019 10:46:54	U	SMPL
PB117246TB - 1	PB117195TB	240	-0.0272	-	7470A	PPB	20 Feb 2019 10:49:00	U	SMPL
K1528-01LX5 - 1		98	-0.0622	-	7470A	PPB	20 Feb 2019 10:51:07	U	SMPL
K1528-01A - 1		18065	4.3636	-	7470A	PPB	20 Feb 2019 10:53:14	U	SMPL
K1525-04LX5 - 1		-9	-0.0885	-	7470A	PPB	20 Feb 2019 10:55:23	U	SMPL
K1525-04A - 1		16821	4.0572	-	7470A	PPB	20 Feb 2019 10:57:30	U	SMPL
CCV90 - 1	CCV90	20850	5.0497	-	7470A	PPB	20 Feb 2019 10:59:37	U	SMPL
CCB90 - 1	CCB90	-35	-0.0949	-	7470A	PPB	20 Feb 2019 11:01:45	U	SMPL