

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB100911

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP70044
calibration std. phosphate 1 ppm	WP71804
calibration std. phosphate 0.5 ppm	WP71803
calibration std. phosphate 0.3 ppm	WP71802
calibration std. 0 ppm	WP71799
phosphate ICV std.	WP71806
phosphate CCV std.	WP71805
calibration std. phosphate 0.1 ppm	WP71801
calibration std. phosphate 0.05 ppm	WP71800
Combined reagent	WP71808
11 N sulfuric acid	WP71598
Phenolphthalein indicator	WP71599
Sodium hydroxide, 1N	WP71600

Intercept: -0.0005Slope: 0.6518Regression: 0.999806

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	02/20/2019	11:30
2	CAL2	0.05	1	50	50	0.034	02/20/2019	11:31
3	CAL3	0.10	1	50	50	0.066	02/20/2019	11:31
4	CAL4	0.30	1	50	50	0.186	02/20/2019	11:32
5	CAL5	0.50	1	50	50	0.331	02/20/2019	11:32
6	CAL6	1.00	1	50	50	0.651	02/20/2019	11:33



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB100911

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.322	0.495	02/20/2019	11:34
2	ICB		1	50	50	0.000	0.001	02/20/2019	11:34
3	CCV1	0.50	1	50	50	0.327	0.502	02/20/2019	11:35
4	CCB1		1	50	50	0.000	0.001	02/20/2019	11:35
5	PB117248BL		1	50	50	0.000	0.001	02/20/2019	11:36
6	PB117248BS	0.50	1	50	50	0.324	0.498	02/20/2019	11:36
7	K1421-08		1	50	50	0.073	0.113	02/20/2019	11:37
8	K1421-08DUP		1	50	50	0.066	0.102	02/20/2019	11:37
9	K1421-08MS	0.50	1	50	50	0.369	0.567	02/20/2019	11:38
10	K1421-08MSD	0.50	1	50	50	0.363	0.558	02/20/2019	11:38
11	CCV2	0.50	1	50	50	0.325	0.499	02/20/2019	11:39
12	CCB2		1	50	50	0.000	0.001	02/20/2019	11:39