


LB100914
INSTRUMENT ID : CV1

7471B

Linear

μ Abs.:

57298



A= 0.0000e+000

B= 1.7216e-004

C= -7.3527e-002

Rho= 0.9993821

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.047	-0.047	153	0.000	153				
0.2	0.200	0.176	-0.024	1448	0.0 %	1448				
2.5	2.500	2.470	-0.030	14773	0.0 %	14773				
5.0	5.000	5.104	0.104	30074	0.0 %	30074				
7.5	7.500	7.707	0.207	45191	0.0 %	45191				
10.0	10.000	9.791	-0.209	57298	0.0 %	57298				

LB100914 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 20 Feb 2019 11:48:54

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	153	-	0.0000	7471B	PPB	20 Feb 2019 11:49:44	S	Std
0.2 - 1	50-2	1448	-	0.2000	7471B	PPB	20 Feb 2019 11:52:02	S	Std
2.5 - 1	52-5	14773	-	2.5000	7471B	PPB	20 Feb 2019 11:54:20	S	Std
5.0 - 1	55-0	30074	-	5.0000	7471B	PPB	20 Feb 2019 11:56:37	S	Std
7.5 - 1	57-5	45191	-	7.5000	7471B	PPB	20 Feb 2019 12:06:01	S	Std
10.0 - 1	510-0	57298	-	10.0000	7471B	PPB	20 Feb 2019 12:11:33	S	Std
ICV18 - 1	ICV18	22577	3.8134	-	7471B	PPB	20 Feb 2019 12:16:10	U	SMPL
ICB18 - 1	ICB18	84	-0.0591	-	7471B	PPB	20 Feb 2019 12:18:25	U	SMPL
CCV91 - 1	CCV91	30108	5.1099	-	7471B	PPB	20 Feb 2019 12:23:17	U	SMPL
CCB91 - 1	CCB91	43	-0.0661	-	7471B	PPB	20 Feb 2019 12:29:02	U	SMPL
CRA - 1	CRA	1479	0.1811	-	7471B	PPB	20 Feb 2019 12:45:12	U	SMPL
HighStd - 1	HighStd	57348	9.7996	-	7471B	PPB	20 Feb 2019 12:47:27	U	SMPL
ChkStd - 1	ChkStd	41014	6.9875	-	7471B	PPB	20 Feb 2019 12:49:43	U	SMPL
PB117225BL - 1	PBS	218	-0.0360	-	7471B	PPB	20 Feb 2019 12:52:00	U	SMPL
PB117225BS - 1	LCSS	22456	3.7925	-	7471B	PPB	20 Feb 2019 12:57:33	U	SMPL
K1349-09 - 1	OR-02-021919-A	3756	0.5731	-	7471B	PPB	20 Feb 2019 12:59:49	U	SMPL
K1349-11 - 1	OR-02-021919-B	3793	0.5795	-	7471B	PPB	20 Feb 2019 13:02:07	U	SMPL
K1525-03 - 1	M-14-S	1370	0.1623	-	7471B	PPB	20 Feb 2019 13:04:24	U	SMPL
K1525-03DUP - 1	M-14-SDUP	1391	0.1659	-	7471B	PPB	20 Feb 2019 13:06:41	U	SMPL
K1525-03MS - 1	M-14-SMS	26122	4.4237	-	7471B	PPB	20 Feb 2019 13:11:35	U	SMPL
CCV92 - 1	CCV92	29840	5.0638	-	7471B	PPB	20 Feb 2019 13:13:49	U	SMPL
CCB92 - 1	CCB92	233	-0.0334	-	7471B	PPB	20 Feb 2019 13:16:06	U	SMPL
K1525-03MSD - 1	M-14-SMSD	25476	4.3124	-	7471B	PPB	20 Feb 2019 13:18:22	U	SMPL
K1525-07 - 1	M-14-D	1696	0.2185	-	7471B	PPB	20 Feb 2019 13:20:38	U	SMPL
K1527-01 - 1	SOIL-PILE	2425	0.3440	-	7471B	PPB	20 Feb 2019 13:28:47	U	SMPL
K1525-03LX5 - 1		603	0.0303	-	7471B	PPB	20 Feb 2019 13:31:02	U	SMPL
K1525-03A - 1		25239	4.2716	-	7471B	PPB	20 Feb 2019 13:33:18	U	SMPL
CCV93 - 1	CCV93	29933	5.0798	-	7471B	PPB	20 Feb 2019 13:35:35	U	SMPL
CCB93 - 1	CCB93	264	-0.0281	-	7471B	PPB	20 Feb 2019 13:37:52	U	SMPL