

LB100921


ISM02.4

INSTRUMENT ID : CV1

Linear

μ Abs.:

4994E



A= 0.0000e+000

B= 2.0106e-004

C= -1.9775e-002

Rho= 0.9999903

Accept=Accepted

*dope
y intercept*

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	-0.011	-0.011	43	0.000	43				
0.05	0.050	0.064	0.014	415	0.0 %	415				
0.2	0.200	0.214	0.014	1164	0.0 %	1164				
2.5	2.500	2.487	-0.013	12466	0.0 %	12466				
5.0	5.000	5.001	0.001	24971	0.0 %	24971				
7.5	7.500	7.474	-0.026	37271	0.0 %	37271				
10	10.000	10.022	0.022	49946	0.0 %	49946				

*% D
28
7
0000*

INSTRUMENT ID: CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0	43	-		0 ISM02.4	PPB	2/20/2019 16:09	Std
	0.05	415	-	0.05	ISM02.4	PPB	2/20/2019 16:12	Std
	0.2	1164	-	0.2	ISM02.4	PPB	2/20/2019 16:14	Std
	2.5	12466	-	2.5	ISM02.4	PPB	2/20/2019 16:16	Std
	5	24971	-	5	ISM02.4	PPB	2/20/2019 16:19	Std
	7.5	37271	-	7.5	ISM02.4	PPB	2/20/2019 16:21	Std
10/1/		49946	-	10	ISM02.4	PPB	2/20/2019 16:23	Std
ICV	ICV	19291	3.8588 -		ISM02.4	PPB	2/20/2019 16:27	SMPL
ICB	ICB	66	-0.0065 -		ISM02.4	PPB	2/20/2019 16:29	SMPL
CCV094	CCV094	24990	5.0046 -		ISM02.4	PPB	2/20/2019 16:31	SMPL
CCB094	CCB094	-31	-0.026 -		ISM02.4	PPB	2/20/2019 16:33	SMPL
PB117239BL	PBS239	188	0.018 -		ISM02.4	PPB	2/20/2019 16:36	SMPL
K1456-01	MBE801	846007	HIGH -		ISM02.4	PPB	2/20/2019 16:38	
K1456-02	MBE801D	879968	HIGH -		ISM02.4	PPB	2/20/2019 16:40	
K1456-03	MBE801S	791160	HIGH -		ISM02.4	PPB	2/20/2019 16:42	
K1456-01DLX10		192078	38.5985 -		ISM02.4	PPB	2/20/2019 16:44	SMPL
K1456-02DLX10		161929	32.5369 -		ISM02.4	PPB	2/20/2019 16:46	SMPL
K1456-03DLX10		164190	32.9915 -		ISM02.4	PPB	2/20/2019 16:48	SMPL
K1456-01DLX50		40398	8.1024 -		ISM02.4	PPB	2/20/2019 16:54	SMPL
K1456-02DLX50		33834	6.7827 -		ISM02.4	PPB	2/20/2019 16:56	SMPL
K1456-03DLX50		34075	6.8312 -		ISM02.4	PPB	2/20/2019 16:58	SMPL
CCV095	CCV095	24834	4.9732 -		ISM02.4	PPB	2/20/2019 17:00	SMPL
CCB095	CCB095	-53	-0.0304 -		ISM02.4	PPB	2/20/2019 17:03	SMPL