



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB100923

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	02/20/2019	15:35
2	CAL2	1	Water	NA	NA	21.1	7.01	02/20/2019	15:36
3	CAL3	1	Water	NA	NA	21.1	10.02	02/20/2019	15:37
4	ICV	1	Water	NA	NA	21.2	6.99	02/20/2019	15:38
5	CCV1	1	Water	NA	NA	21.1	2.02	02/20/2019	15:39
6	K1544-04	1	Water	NA	NA	12.1	5.87	02/20/2019	15:40
7	K1544-05	1	Water	NA	NA	10.2	5.35	02/20/2019	15:42
8	K1544-06	1	Water	NA	NA	10.8	5.62	02/20/2019	15:43
9	K1544-06DUP	1	Water	NA	NA	10.9	5.63	02/20/2019	15:44
10	CCV2	1	Water	NA	NA	21.1	12.01	02/20/2019	15:45

7B 100923

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K1344

WorkList ID : 122520

Date : 02-20-2019 14:06:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/20/2019	Water	K1544-04 - 1A	pH	Cool 4 deg C	PSEG01	K31	COMP-1	02/20/2019	9040C
02/20/2019	Water	K1544-05 - 1A	pH	Cool 4 deg C	PSEG01	K31	21803	02/20/2019	9040C
02/20/2019	Water	K1544-06 - 1A	pH	Cool 4 deg C	PSEG01	K31	COMP-2	02/20/2019	9040C

Date/Time 02-20-19 15:30
Received by: JD
Relinquished by: CP

Date/Time 02-20-19 16:10
Received by: JD
Relinquished by: CP