



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB100924

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	02/20/2019	15:35
2	CAL2	1	Water	NA	NA	21.1	7.01	02/20/2019	15:36
3	CAL3	1	Water	NA	NA	21.1	10.02	02/20/2019	15:37
4	ICV	1	Water	NA	NA	21.2	6.99	02/20/2019	15:38
5	CCV1	1	Water	NA	NA	21.1	2.02	02/20/2019	16:15
6	K1549-04	1	Water	NA	NA	16.8	7.60	02/20/2019	16:16
7	K1550-03	1	Water	NA	NA	14.3	6.24	02/20/2019	16:17
8	K1550-03DUP	1	Water	NA	NA	14.4	6.25	02/20/2019	16:18
9	CCV2	1	Water	NA	NA	21.1	12.01	02/20/2019	16:20

2B 100924

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K1351

WorkList ID : 122523

Date : 02-20-2019 16:01:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/20/2019	Water	K1549-04 -A	Corrosivity	Cool 4 deg C	PSEG01	K31	NB-022019	02/20/2019	9040C
02/20/2019	Water	K1550-03 r1	Corrosivity	Cool 4 deg C	PSEG01	K11	SS022019	02/20/2019	9040C

Date/Time 02-20-19 16:10
Received by: JD
Relinquished by: CO

Date/Time 02-20-19 16:35
Received by: JD
Relinquished by: CO