



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: Non-Polar Material
Run Number: LB100931
Analysis Date: 02/21/2019
BalanceID: WC SC-3
OvenID: EX OVEN

ANALYST: jignesh
REVIEWED BY: apatel
Extraction Date: 02/21/2019
Extraction IN Time: 08:10
Extraction OUT Time: 08:45

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB100931BL	LB100931BL	WATER	2	1000	100	2.4033	2.4033	3.02	2.4033	2.4033	0.0000	0
2	LB100931BS	LB100931BS	WATER	2	1000	100	2.8477	2.8477	3.02	2.8668	2.8668	0.0191	19.1
3	LB100931BSD	LB100931BSD	WATER	2	1000	100	2.8401	2.8401	3.01	2.8593	2.8593	0.0192	19.2
4	K1545-01	NON-POLAR-MATERIAL	WATER	2	100	100	3.2691	3.2691	3.04	3.2699	3.2699	0.0008	8
5	K1545-02	NON-POLAR-MATERIAL	WATER	2	100	100	3.0474	3.0474	3.03	3.0481	3.0481	0.0007	7
6	K1545-03	NON-POLAR-MATERIAL	WATER	2	100	100	3.2337	3.2337	3.04	3.2343	3.2343	0.0006	6
7	K1545-04	NON-POLAR-MATERIAL	WATER	2	100	100	3.0384	3.0384	3.02	3.0390	3.0390	0.0006	6



QC Batch# LB100931 Test: Non-Polar Ma Analysis Date: 02/21/2019

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2478
pH Paper 0-14	W2435
Sodium Sulfate	EP1922
1:1 HCL	WP69312
Silica Gel	W2466
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP71442
LCSWD	10.00 ML	WP71443
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance:	<u>0.0019</u>	In OVEN TEMP :	<u>70</u>	Dessicator Time In :	<u>10:16</u>
1.0000 gram Balance:	<u>1.0004</u>	In Time:	<u>09:15</u>		
Bal Check Time:	<u>08:30</u>	Out OVEN TEMP:	<u>70</u>	Dessicator Time Out:	<u>10:35</u>
		Out Time:	<u>10:15</u>		

After Analysis

0.0020 gram Balance:	<u>0.0021</u>	In OVEN TEMP :	<u>71</u>	Dessicator Time In :	<u>11:31</u>
1.0000 gram Balance:	<u>1.0005</u>	In Time:	<u>11:00</u>		
Bal Check Time:	<u>12:35</u>	Out OVEN TEMP:	<u>71</u>	Dessicator Time Out:	<u>12:05</u>
		Out Time:	<u>11:30</u>		

LB 100931

WORKLIST(Hardcopy Internal Chain)

WorkList Name : NON POLAR K1545

WorkList ID : 122528

Date : 02-21-2019 07:51:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/05/2019	Water	K1545-01 ✓	Non-Polar Material	Conc H2SO4 to	AMER08	M63	NON-POLAR-MATERIAL	02/20/2019	1664A
03/05/2019	Water	K1545-02 ✓	Non-Polar Material	Conc H2SO4 to	AMER08	M63	NON-POLAR-MATERIAL	02/20/2019	1664A
03/05/2019	Water	K1545-03 ✓	Non-Polar Material	Conc H2SO4 to	AMER08	M63	NON-POLAR-MATERIAL	02/20/2019	1664A
03/05/2019	Water	K1545-04 ✓	Non-Polar Material	Conc H2SO4 to	AMER08	M63	NON-POLAR-MATERIAL	02/20/2019	1664A

Date/Time 02-21-19 8:00 AM

Received by: JB

Relinquished by: 

Date/Time 02-21-19 9:45

Received by: 

Relinquished by: JB