



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB100946

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.0	80.6	80.9	02/21/2019	09:35
2	K1544-04		1	7	>100.0	>212.0	>212.0	02/21/2019	09:55
3	K1544-04DUP		1	7	>100.0	>212.0	>212.0	02/21/2019	10:20
4	K1544-05		1	9	>100.0	>212.0	>212.0	02/21/2019	10:45
5	K1544-06		1	6	>100.0	>212.0	>212.0	02/21/2019	11:10
6	K1549-04		1	7	>100.0	>212.0	>212.0	02/21/2019	11:40
7	K1550-03		1	8	>100.0	>212.0	>212.0	02/21/2019	12:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB100946

WorkList Name : FP22119

WorkList ID : 122553

Date : 02-21-2019 09:08:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/27/2019	Water	K1544-04	F	Cool 4 deg C	PSEG01	K31	COMP-1	02/20/2019	1010A
02/27/2019	Water	K1544-05	F	Cool 4 deg C	PSEG01	K31	21803	02/20/2019	1010A
02/27/2019	Water	K1544-06	F	Cool 4 deg C	PSEG01	K31	COMP-2	02/20/2019	1010A
02/25/2019	Water	K1549-04	E	Cool 4 deg C	PSEG01	K31	NB-022019	02/20/2019	1010A
02/25/2019	Water	K1550-03	G	Cool 4 deg C	PSEG01	K11	SS022019	02/20/2019	1010A

Date/Time 02/21/19 9:10
 Received by: KS
 Relinquished by: CP

Date/Time 02/21/19 12:10
 Received by: CP
 Relinquished by: KS