

LB100950


ISM02.4

INSTRUMENT ID: CV1

Linear

 μ Abs.:

44059



A= 0.0000e+000

B= 2.2727e-004

C= 1.0985e-003

Rho= 0.9999438

Accept=Accepted

Slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.001	-0.001	-8	0.000	-8					-5
0.2	0.200	0.190	-0.010	831	0.0 %	831					2
2.5	2.500	2.561	0.061	11262	0.0 %	11262					-1
5.0	5.000	4.928	-0.072	21680	0.0 %	21680					0
7.5	7.500	7.508	0.008	33029	0.0 %	33029					0
10	10.000	10.014	0.014	44059	0.0 %	44059					0

LB100950

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	-8	-		0 ISM02.4	PPB	2/22/2019 9:18	Std
	0.2 S0.2	831	-		0.2 ISM02.4	PPB	2/22/2019 9:21	Std
	2.5 S2.5	11262	-		2.5 ISM02.4	PPB	2/22/2019 9:23	Std
	5 S5.0	21680	-		5 ISM02.4	PPB	2/22/2019 9:25	Std
	7.5 S7.5	33029	-		7.5 ISM02.4	PPB	2/22/2019 9:27	Std
	10 S10.0	44059	-		10 ISM02.4	PPB	2/22/2019 9:30	Std
ICV	ICV	17630	4.0251 -		ISM02.4	PPB	2/22/2019 9:33	SMPL
ICB	ICB	-19	0.0056 -		ISM02.4	PPB	2/22/2019 9:35	SMPL
CCV001	CCV001	21903	4.9983 -		ISM02.4	PPB	2/22/2019 9:37	SMPL
CCB001	CCB001	-6	0.0085 -		ISM02.4	PPB	2/22/2019 9:40	SMPL
PB117288BL	PBW288	110	0.0349 -		ISM02.4	PPB	2/22/2019 9:42	SMPL
K1557-01	MC0AG8	80	0.0281 -		ISM02.4	PPB	2/22/2019 9:44	SMPL
K1557-02	MC0AG9	225	0.0611 -		ISM02.4	PPB	2/22/2019 9:46	SMPL
K1557-03	MC0AG9D	238	0.0641 -		ISM02.4	PPB	2/22/2019 9:49	SMPL
K1557-04	MC0AG9S	4317	0.9931 -		ISM02.4	PPB	2/22/2019 9:51	SMPL
K1557-05	MC0AH0	185	0.052 -		ISM02.4	PPB	2/22/2019 9:53	SMPL
K1557-06	MC0AH1	84	0.029 -		ISM02.4	PPB	2/22/2019 9:55	SMPL
K1557-07	MC0AH2	74	0.0267 -		ISM02.4	PPB	2/22/2019 9:58	SMPL
K1557-08	MC0AH3	130	0.0395 -		ISM02.4	PPB	2/22/2019 10:00	SMPL
CCV002	CCV002	21368	4.8764 -		ISM02.4	PPB	2/22/2019 10:02	SMPL
CCB002	CCB002	-20	0.0053 -		ISM02.4	PPB	2/22/2019 10:05	SMPL