



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB100952

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	02/21/2019	15:10
2	CAL2	1	Water	NA	NA	21.2	7.00	02/21/2019	15:11
3	CAL3	1	Water	NA	NA	21.2	10.02	02/21/2019	15:12
4	ICV	1	Water	NA	NA	21.3	7.01	02/21/2019	15:14
5	CCV1	1	Water	NA	NA	21.2	2.01	02/21/2019	15:15
6	K1556-02	1	Solid	20.02	20	21.8	6.67	02/21/2019	15:16
7	K1556-02DUP	1	Solid	20.03	20	21.9	6.68	02/21/2019	15:17
8	CCV2	1	Water	NA	NA	21.2	12.01	02/21/2019	15:20

LB 100952

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K1556

WorkList ID : 122569

Date : 02-21-2019 15:51:12

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/21/2019	Solid	K1556-02	Corrosivity	Cool 4 deg C	PSEG01	K31	COMP	02/21/2019	9045D

Date/Time 02-21-19 3:00 PM
 Received by: J
 Relinquished by: 26

Date/Time 02-21-19 16:00
 Received by: DP
 Relinquished by: 28