

LB100964


INSTRUMENT ID: C01

7470A

Linear

 μ Abs.:

42933



A= 0.0000e+000

B= 2.3321e-004

C= -3.8080e-002

Rho= 0.9998799

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.033	-0.033	22	0.000	22				
0.2	0.200	0.157	-0.043	836	0.0 %	836				
2.5	2.500	2.552	0.052	11107	0.0 %	11107				
5.0	5.000	5.103	0.103	22043	0.0 %	22043				
7.5	7.500	7.447	-0.053	32097	0.0 %	32097				
10.0	10.000	9.974	-0.026	42933	0.0 %	42933				

LB100964 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 Feb 2019 10:42:29

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	22	-	0.0000	7470A	PPB	22 Feb 2019 10:42:38	S	Std
0.2 - 1	50-2	836	-	0.2000	7470A	PPB	22 Feb 2019 10:44:45	S	Std
2.5 - 1	52-5	11107	-	2.5000	7470A	PPB	22 Feb 2019 10:46:53	S	Std
5.0 - 1	55-0	22043	-	5.0000	7470A	PPB	22 Feb 2019 10:49:01	S	Std
7.5 - 1	57-5	32097	-	7.5000	7470A	PPB	22 Feb 2019 10:51:09	S	Std
10.0 - 1	510-0	42933	-	10.0000	7470A	PPB	22 Feb 2019 10:55:45	S	Std
ICV20 - 1	ICV20	16954	3.9157	-	7470A	PPB	22 Feb 2019 10:59:42	U	SMPL
ICB20 - 1	ICB20	-32	-0.0455	-	7470A	PPB	22 Feb 2019 11:01:48	U	SMPL
CCV03 - 1	CCV03	20945	4.8465	-	7470A	PPB	22 Feb 2019 11:03:54	U	SMPL
CCB03 - 1	CCB03	-27	-0.0444	-	7470A	PPB	22 Feb 2019 11:06:00	U	SMPL
CRA - 1	CRA	990	0.1928	-	7470A	PPB	22 Feb 2019 11:08:06	U	SMPL
HighStd - 1	HighStd	43970	10.2161	-	7470A	PPB	22 Feb 2019 11:10:12	U	SMPL
ChkStd - 1	ChkStd	29488	6.8387	-	7470A	PPB	22 Feb 2019 11:12:17	U	SMPL
PB117281BL - 1	PBW	61	-0.0239	-	7470A	PPB	22 Feb 2019 11:14:25	U	SMPL
PB117281BS - 1	LCSW	16879	3.8982	-	7470A	PPB	22 Feb 2019 11:16:32	U	SMPL
K1544-04 - 1	COMP-1	151	-0.0029	-	7470A	PPB	22 Feb 2019 11:18:39	U	SMPL
K1544-05 - 1	21803	192	0.0067	-	7470A	PPB	22 Feb 2019 11:20:46	U	SMPL
K1544-06 - 1	COMP-2	157	-0.0015	-	7470A	PPB	22 Feb 2019 11:22:54	U	SMPL
K1544-06DUP - 1	COMP-2DUP	162	-0.0003	-	7470A	PPB	22 Feb 2019 11:25:02	U	SMPL
K1544-06MS - 1	COMP-2MS	17230	3.9801	-	7470A	PPB	22 Feb 2019 11:27:09	U	SMPL
CCV04 - 1	CCV04	22065	5.1076	-	7470A	PPB	22 Feb 2019 11:29:15	U	SMPL
CCB04 - 1	CCB04	-22	-0.0432	-	7470A	PPB	22 Feb 2019 11:31:21	U	SMPL
K1544-06MSD - 1	COMP-2MSD	17257	3.9864	-	7470A	PPB	22 Feb 2019 11:33:27	U	SMPL
PB117282BL - 1	PBW	114	-0.0115	-	7470A	PPB	22 Feb 2019 11:35:34	U	SMPL
PB117282BS - 1	LCSW	16291	3.7611	-	7470A	PPB	22 Feb 2019 11:37:40	U	SMPL
K1344-02 - 1	NB-08-022019-A	208	0.0104	-	7470A	PPB	22 Feb 2019 11:39:47	U	SMPL
K1344-04 - 1	NB-08-022019-B	306	0.0333	-	7470A	PPB	22 Feb 2019 11:41:53	U	SMPL
K1344-06 - 1	NB-07-022019-A	79	-0.0197	-	7470A	PPB	22 Feb 2019 11:44:00	U	SMPL
K1344-08 - 1	NB-07-022019-B	311	0.0344	-	7470A	PPB	22 Feb 2019 11:46:07	U	SMPL
K1345-04 - 1	HD-01-022019	182	0.0044	-	7470A	PPB	22 Feb 2019 11:48:15	U	SMPL
K1349-10 - 1	OR-02-021919-A	208	0.0104	-	7470A	PPB	22 Feb 2019 11:50:23	U	SMPL
K1349-10DUP - 1	OR-02-021919-ADUP	202	0.0090	-	7470A	PPB	22 Feb 2019 11:52:30	U	SMPL
CCV05 - 1	CCV05	21942	5.0790	-	7470A	PPB	22 Feb 2019 11:54:36	U	SMPL
CCB05 - 1	CCB05	-24	-0.0437	-	7470A	PPB	22 Feb 2019 11:56:43	U	SMPL
K1349-10MS - 1	OR-02-021919-AMS	16598	3.8327	-	7470A	PPB	22 Feb 2019 11:58:50	U	SMPL
K1349-10MSD - 1	OR-02-021919-AMSD	16369	3.7793	-	7470A	PPB	22 Feb 2019 12:00:56	U	SMPL
K1349-12 - 1	OR-02-021919-B	120	-0.0101	-	7470A	PPB	22 Feb 2019 12:03:03	U	SMPL
K1527-03 - 1	SOIL-PILE	155	-0.0019	-	7470A	PPB	22 Feb 2019 12:05:09	U	SMPL
K1537-04 - 1	DRUM-1-2	171	0.0018	-	7470A	PPB	22 Feb 2019 12:07:15	U	SMPL
K1538-04 - 1	M-11-A-S	157	-0.0015	-	7470A	PPB	22 Feb 2019 12:09:23	U	SMPL
K1538-08 - 1	M-11-A-D	146	-0.0040	-	7470A	PPB	22 Feb 2019 12:11:30	U	SMPL
K1544-01 - 1	21804	161	-0.0005	-	7470A	PPB	22 Feb 2019 12:13:38	U	SMPL
K1544-02 - 1	21805	147	-0.0038	-	7470A	PPB	22 Feb 2019 12:15:45	U	SMPL
K1544-03 - 1	CHRT-20420	131	-0.0075	-	7470A	PPB	22 Feb 2019 12:17:52	U	SMPL
CCV06 - 1	CCV06	21664	5.0141	-	7470A	PPB	22 Feb 2019 12:19:59	U	SMPL
CCB06 - 1	CCB06	70	-0.0218	-	7470A	PPB	22 Feb 2019 12:22:05	U	SMPL
K1546-03 - 1	RBR-200036	127	-0.0085	-	7470A	PPB	22 Feb 2019 12:24:12	U	SMPL
K1546-06 - 1	RT-3422	132	-0.0073	-	7470A	PPB	22 Feb 2019 12:26:18	U	SMPL
K1546-09 - 1	RBR-200030	149	-0.0033	-	7470A	PPB	22 Feb 2019 12:28:25	U	SMPL
K1546-12 - 1	RBR-200034	147	-0.0038	-	7470A	PPB	22 Feb 2019 12:30:32	U	SMPL
K1546-15 - 1	RT-1495	171	0.0018	-	7470A	PPB	22 Feb 2019 12:32:38	U	SMPL
K1547-04 - 1	TP-1-S	150	-0.0031	-	7470A	PPB	22 Feb 2019 12:34:46	U	SMPL
PB117289BL - 1	PBW	117	-0.0108	-	7470A	PPB	22 Feb 2019 12:36:53	U	SMPL
PB117289BS - 1	LCSW	16852	3.8919	-	7470A	PPB	22 Feb 2019 12:39:01	U	SMPL
K1549-04 - 1	NB-022019	68	-0.0222	-	7470A	PPB	22 Feb 2019 12:41:08	U	SMPL
K1550-03 - 1	SS022019	95	-0.0159	-	7470A	PPB	22 Feb 2019 12:43:15	U	SMPL
CCV07 - 1	CCV07	21320	4.9339	-	7470A	PPB	22 Feb 2019 12:45:21	U	SMPL
CCB07 - 1	CCB07	72	-0.0213	-	7470A	PPB	22 Feb 2019 12:47:28	U	SMPL
K1550-03DUP - 1	SS022019DUP	97	-0.0155	-	7470A	PPB	22 Feb 2019 12:49:35	U	SMPL
K1550-03MS - 1	SS022019MS	14120	3.2548	-	7470A	PPB	22 Feb 2019 12:51:42	U	SMPL
K1550-03MSD - 1	SS022019MSD	13963	3.2182	-	7470A	PPB	22 Feb 2019 12:53:49	U	SMPL
PB117281TB - 1	PB117255TB	109	-0.0127	-	7470A	PPB	22 Feb 2019 13:00:22	U	SMPL
PB117282TB - 1	PB117249TB	176	0.0030	-	7470A	PPB	22 Feb 2019 13:02:29	U	SMPL
K1544-06LX5 - 1		168	0.0011	-	7470A	PPB	22 Feb 2019 13:04:37	U	SMPL
K1544-06A - 1		17416	4.0235	-	7470A	PPB	22 Feb 2019 13:06:44	U	SMPL
K1349-10LX5 - 1		109	-0.0127	-	7470A	PPB	22 Feb 2019 13:08:51	U	SMPL
K1349-10A - 1		16440	3.7959	-	7470A	PPB	22 Feb 2019 13:10:59	U	SMPL
K15150-03LX5 - 1		78	-0.0199	-	7470A	PPB	22 Feb 2019 13:13:07	U	SMPL
CCV08 - 1	CCV08	21934	5.0771	-	7470A	PPB	22 Feb 2019 13:15:15	U	SMPL
CCB08 - 1	CCB08	151	-0.0029	-	7470A	PPB	22 Feb 2019 13:17:22	U	SMPL
K15150-03A - 1		14706	3.3915	-	7470A	PPB	22 Feb 2019 13:19:28	U	SMPL
CCV09 - 1	CCV09	21684	5.0188	-	7470A	PPB	22 Feb 2019 13:21:35	U	SMPL
CCB09 - 1	CCB09	78	-0.0199	-	7470A	PPB	22 Feb 2019 13:23:42	U	SMPL
					7470A	PPB	22 Feb 2019 13:25:49	U	SMPL