



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB100965

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.1	4.01	02/21/2019	14:25
2	CAL2	1	Water	NA	NA	22.1	7.01	02/21/2019	14:26
3	CAL3	1	Water	NA	NA	22.1	10.02	02/21/2019	14:27
4	ICV	1	Water	NA	NA	22.1	7.00	02/21/2019	14:30
5	CCV1	1	Water	NA	NA	22.1	2.02	02/21/2019	14:31
6	K1544-01	1	Solid	20.03	20	21.2	5.10	02/21/2019	14:32
7	K1544-02	1	Solid	20.04	20	21.5	5.80	02/21/2019	14:33
8	K1544-03	1	Solid	20.02	20	20.3	6.05	02/21/2019	14:34
9	K1544-03DUP	1	Solid	20.03	20	20.4	6.04	02/21/2019	14:35
10	CCV2	1	Water	NA	NA	22.2	12.01	02/21/2019	14:36

23 100965

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K1544 WorkList ID : 122611 Date : 02-21-2019 14:15:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/20/2019	Solid	K1544-01	pH	Cool 4 deg C	PSEG01	K31	21804	02/20/2019	9045D
02/20/2019	Solid	K1544-02	pH	Cool 4 deg C	PSEG01	K31	21805	02/20/2019	9045D
02/20/2019	Solid	K1544-03	pH	Cool 4 deg C	PSEG01	K31	CHRT-20420	02/20/2019	9045D

Date/Time 02-21-19 14:20
 Received by: JS
 Relinquished by: CS

Date/Time 02-21-19 15:15
 Received by: JS
 Relinquished by: CS