



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB100968

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	02/22/2019	12:15
2	CAL2	1	Water	NA	NA	21.1	7.01	02/22/2019	12:16
3	CAL3	1	Water	NA	NA	21.1	10.02	02/22/2019	12:17
4	ICV	1	Water	NA	NA	21.2	6.99	02/22/2019	12:18
5	CCV1	1	Water	NA	NA	21.2	2.01	02/22/2019	12:20
6	K1591-01	1	Water	NA	NA	18.0	5.57	02/22/2019	12:21
7	K1591-01DUP	1	Water	NA	NA	18.1	5.58	02/22/2019	12:22
8	CCV2	1	Water	NA	NA	21.1	12.01	02/22/2019	12:24

2B 100968

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K1591

WorkList ID : 122613

Date : 02-22-2019 12:02:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/21/2019	Water	K1591-01-A	pH	Cool 4 deg C	STER01	M51	POOL-WATER	02/21/2019	SM 4500-H B

Date/Time 02-22-19 12:05
Received by: J
Relinquished by: CO

Date/Time 02-22-19 12:50
Received by: 30
Relinquished by: CJ