

LB100974


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

42972



A= 0.0000e+000

B= 2.3217e-004

C= 2.1946e-002

Rho= 0.9999292

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.004	0.004	-79	0.000	-79				
0.2	0.200	0.249	0.049	979	0.0 %	979				
2.5	2.500	2.457	-0.043	10486	0.0 %	10486				
5.0	5.000	4.936	-0.064	21165	0.0 %	21165				
7.5	7.500	7.556	0.056	32450	0.0 %	32450				
10.0	10.000	9.999	-0.001	42972	0.0 %	42972				

LB100974 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 Feb 2019 14:28:29

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	-79	-	0.0000	7470A	PPB	22 Feb 2019 15:40:14	S	Std
0.2 - 1	50.2	979	-	0.2000	7470A	PPB	22 Feb 2019 15:42:22	S	Std
2.5 - 1	52.5	10486	-	2.5000	7470A	PPB	22 Feb 2019 15:44:30	S	Std
5.0 - 1	55.0	21165	-	5.0000	7470A	PPB	22 Feb 2019 15:46:38	S	Std
7.5 - 1	57.5	32450	-	7.5000	7470A	PPB	22 Feb 2019 15:48:46	S	Std
10.0 - 1	510.0	42972	-	10.0000	7470A	PPB	22 Feb 2019 15:50:53	S	Std
ICV21 - 1	ICV21	16654	3.8885	-	7470A	PPB	22 Feb 2019 15:55:51	U	SMPL
ICB21 - 1	ICB21	-43	0.0120	-	7470A	PPB	22 Feb 2019 15:57:57	U	SMPL
CCV10 - 1	CCV10	21568	5.0294	-	7470A	PPB	22 Feb 2019 16:00:03	U	SMPL
CCB10 - 1	CCB10	-11	0.0194	-	7470A	PPB	22 Feb 2019 16:02:09	U	SMPL
CRA - 1	CRA	843	0.2177	-	7470A	PPB	22 Feb 2019 16:20:59	U	SMPL
HighStd - 1	HighStd	43687	10.1649	-	7470A	PPB	22 Feb 2019 16:23:04	U	SMPL
ChkStd - 1	ChkStd	29966	6.9792	-	7470A	PPB	22 Feb 2019 16:25:11	U	SMPL
PB117323BL - 1	PBW	-24	0.0164	-	7470A	PPB	22 Feb 2019 16:27:18	U	SMPL
PB117323BS - 1	LCSW	16075	3.7541	-	7470A	PPB	22 Feb 2019 16:29:26	U	SMPL
K1536-01 - 1	OILY-STONES	3	0.0226	-	7470A	PPB	22 Feb 2019 16:31:33	U	SMPL
K1556-02 - 1	COMP	32	0.0294	-	7470A	PPB	22 Feb 2019 16:33:40	U	SMPL
K1556-02DUP - 1	COMPDUP	33	0.0296	-	7470A	PPB	22 Feb 2019 16:35:48	U	SMPL
K1556-02MS - 1	COMPMS	16580	3.8714	-	7470A	PPB	22 Feb 2019 16:37:56	U	SMPL
K1556-02MSD - 1	COMPMSD	16457	3.8428	-	7470A	PPB	22 Feb 2019 16:40:01	U	SMPL
CCV11 - 1	CCV11	21207	4.9456	-	7470A	PPB	22 Feb 2019 16:42:07	U	SMPL
CCB11 - 1	CCB11	14	0.0252	-	7470A	PPB	22 Feb 2019 16:44:14	U	SMPL
K1564-01 - 1	02-21-DRUM-A	19	0.0264	-	7470A	PPB	22 Feb 2019 16:46:25	U	SMPL
K1564-02 - 1	02-21-DRUM-B	35	0.0301	-	7470A	PPB	22 Feb 2019 16:48:31	U	SMPL
K1564-03 - 1	02-21-DRUM-C	22	0.0271	-	7470A	PPB	22 Feb 2019 16:50:37	U	SMPL
K1564-04 - 1	02-21-DRUM-D	26	0.0280	-	7470A	PPB	22 Feb 2019 16:52:44	U	SMPL
K1564-05 - 1	02-21-DRUM-E	32	0.0294	-	7470A	PPB	22 Feb 2019 16:54:50	U	SMPL
K1564-06 - 1	02-21-DRUM-F	45	0.0324	-	7470A	PPB	22 Feb 2019 16:56:58	U	SMPL
K1564-07 - 1	02-21-DRUM-G	27	0.0282	-	7470A	PPB	22 Feb 2019 16:59:05	U	SMPL
PB117324BL - 1	PBW	53	0.0343	-	7470A	PPB	22 Feb 2019 17:01:13	U	SMPL
PB117324BS - 1	LCSW	15169	3.5438	-	7470A	PPB	22 Feb 2019 17:03:21	U	SMPL
K1561-01 - 1	23719	3	0.0226	-	7470A	PPB	22 Feb 2019 17:05:28	U	SMPL
CCV12 - 1	CCV12	21230	4.9510	-	7470A	PPB	22 Feb 2019 17:07:34	U	SMPL
CCB12 - 1	CCB12	116	0.0489	-	7470A	PPB	22 Feb 2019 17:09:41	U	SMPL
K1572-03 - 1	COMP-1	74	0.0391	-	7470A	PPB	22 Feb 2019 17:11:47	U	SMPL
K1572-03DUP - 1	COMP-1DUP	36	0.0303	-	7470A	PPB	22 Feb 2019 17:13:54	U	SMPL
K1572-03MS - 1	COMP-1MS	16553	3.8651	-	7470A	PPB	22 Feb 2019 17:16:00	U	SMPL
K1572-03MSD - 1	COMP-1MSD	16582	3.8718	-	7470A	PPB	22 Feb 2019 17:18:07	U	SMPL
PB117323TB - 1	PB117285TB	18	0.0261	-	7470A	PPB	22 Feb 2019 17:20:14	U	SMPL
PB117324TB - 1	PB117290TB	59	0.0356	-	7470A	PPB	22 Feb 2019 17:22:21	U	SMPL
K1556-02LX5 - 1		23	0.0273	-	7470A	PPB	22 Feb 2019 17:24:29	U	SMPL
K1556-02A - 1		16437	3.8382	-	7470A	PPB	22 Feb 2019 17:26:37	U	SMPL
K1572-03LX5 - 1		7	0.0236	-	7470A	PPB	22 Feb 2019 17:28:44	U	SMPL
K1572-03A - 1		16945	3.9561	-	7470A	PPB	22 Feb 2019 17:30:51	U	SMPL
CCV13 - 1	CCV13	21108	4.9226	-	7470A	PPB	22 Feb 2019 17:32:57	U	SMPL
CCB13 - 1	CCB13	-56	0.0089	-	7470A	PPB	22 Feb 2019 17:35:03	U	SMPL