



PERCENT SOLID

Supervisor: ketankumar
Analyst: jignesh
Date: 2/23/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:25
In Date: 02/22/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:20
Out Date: 02/23/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB100976

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K1348-03	OK-03-022219-A	1	1.13	9.9	8.93	88.9
K1348-05	OK-03-022219-B	2	1.18	9.97	9.41	93.6
K1356-09	SU-02-022219-A	3	1.15	9.7	8.42	85.0
K1356-11	SU-02-022219-B	4	1.15	9.94	8.84	87.5
K1566-01	23377	5	1.00	2.00	2.00	100.0
K1567-01	COMP-1	6	1.00	2.00	2.00	100.0
K1567-02	9816	7	1.00	2.00	2.00	100.0
K1572-01	COMP-2	8	1.00	2.00	2.00	100.0
K1572-02	30019	9	1.00	2.00	2.00	100.0
K1580-05	SVOC-GPC-BLANK	10	1.00	2.00	2.00	100.0
K1580-06	PEST-GPC-BLANK	11	1.00	2.00	2.00	100.0
K1580-07	PEST-GPC-BLANK-SPIKE	12	1.00	2.00	2.00	100.0
K1580-08	PCB-GPC-BLANK	13	1.00	2.00	2.00	100.0
K1580-09	PCB-GPC-BLANK-SPIKE	14	1.00	2.00	2.00	100.0
K1580-10	SVOC-GPC2-BLANK	15	1.00	2.00	2.00	100.0
K1580-11	PEST-GPC2-BLANK	16	1.00	2.00	2.00	100.0
K1580-12	PEST-GPC2-BLANK-SPIKE	17	1.00	2.00	2.00	100.0
K1580-13	PCB-GPC2-BLANK	18	1.00	2.00	2.00	100.0
K1580-14	PCB-GPC2-BLANK-SPIKE	19	1.00	2.00	2.00	100.0
K1586-01	M-1-1	20	1.16	9.85	8.92	89.3
K1586-02	M-1-2	21	1.14	9.96	8.85	87.4
K1586-03	M-1	22	1.13	9.7	8.67	88.0
K1592-01	TP-4-S1	23	1.16	9.86	8.85	88.4
K1592-02	TP-4-S2	24	1.14	9.81	8.95	90.1
K1592-03	TP-4	25	1.13	9.94	8.7	85.9
K1595-01	TP-CBF-B10	26	1.12	9.71	8.07	80.9
K1595-03	TP-CBF-B11	27	1.13	9.97	8.74	86.1
K1595-05	TP-E-WALL-GRADE-BEAM	28	1.12	9.85	8.29	82.1
K1595-07	TP-MH-C02	29	1.11	9.94	8.63	85.2
K1596-01	COMP-13	30	1.14	9.61	8.35	85.1
K1596-02	COMP-14	31	1.14	9.87	8.64	85.9
K1596-03	COMP-15	32	1.13	9.55	8.17	83.6
K1596-07	VOC-13	33	1.14	5.64	5.16	89.3
K1596-08	VOC-14	34	1.13	6.64	5.96	87.7
K1596-09	VOC-15	35	1.13	5.54	4.99	87.5
K1597-01	TP-25	36	1.14	9.96	8.54	83.9
K1597-03	TP-26	37	1.15	9.98	8.61	84.5
K1597-05	PB-21	38	1.13	9.92	7.8	75.9
K1597-07	PB-20	39	1.15	9.87	8.37	82.8

PERCENT SOLID

Supervisor: ketankumar

Analyst: jignesh

Date: 2/23/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:25
In Date: 02/22/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:20
Out Date: 02/23/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB100976

<u>Lab ID</u>	<u>Client Sample ID</u>	<u>Dish#</u>	<u>Dish Wt(g)</u> <u>(A)</u>	<u>Dish +</u> <u>Sample Wt(g)</u> <u>(B)</u>	<u>Dish + Dry</u> <u>Sample Wt(g)</u> <u>(C)</u>	<u>% Solid</u>
K1597-09	PB-19	40	1.13	9.9	9.44	94.8
K1597-11	TP-22	41	1.13	9.96	8.37	82.0
K1597-13	PB-22	42	1.15	9.97	8.18	79.7
K1597-15	TP-23	43	1.14	9.97	8.8	86.7

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$