



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB100988

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.9	02/23/2019	10:40
2	K1587-01		1	8	>100.0	>212.0	>212.0	02/23/2019	11:05
3	K1587-01DUP		1	8	>100.0	>212.0	>212.0	02/23/2019	11:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 100 988

WorkList Name : FP22319

WorkList ID : 122638

Date : 02-23-2019 09:57:19

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/08/2019	Water	K1587-01	K Flash Point	Cool 4 deg C	TULL02	M41	TP-MH-C02	02/22/2019	1010A

Date/Time 02/23/19 10:05
 Received by: KS
 Relinquished by: KS

Date/Time 02/23/19 11:45
 Received by: KS
 Relinquished by: KS