



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB101001

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	02/22/2019	17:05
2	CAL2	1	Water	NA	NA	21.1	7.01	02/22/2019	17:06
3	CAL3	1	Water	NA	NA	21.1	10.02	02/22/2019	17:07
4	ICV	1	Water	NA	NA	21.1	7.00	02/22/2019	17:08
5	CCV1	1	Water	NA	NA	21.2	2.02	02/22/2019	17:10
6	K1598-01	1	Water	NA	NA	10.2	7.88	02/22/2019	17:11
7	K1598-01DUP	1	Water	NA	NA	10.3	7.89	02/22/2019	17:12
8	CCV2	1	Water	NA	NA	21.1	12.01	02/22/2019	17:13

LB 101001

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K1598

WorkList ID : 122626

Date : 02-22-2019 16:47:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/22/2019	Water	K1598-01 E	pH	Cool 4 deg C	BIOH02	M41	TOTAL-N	02/22/2019	9040C

Date/Time 02.22.19 17:00
Received by: Jd
Relinquished by: CS

Date/Time 02.22.19 17:35
Received by: Jd
Relinquished by: CS