



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB101003

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.1	4.01	02/25/2019	15:20
2	CAL2	1	Water	NA	NA	22.1	7.00	02/25/2019	15:21
3	CAL3	1	Water	NA	NA	22.1	10.02	02/25/2019	15:22
4	ICV	1	Water	NA	NA	22.2	7.01	02/25/2019	15:24
5	CCV1	1	Water	NA	NA	22.1	2.01	02/25/2019	15:25
6	K1605-01	1	Water	NA	NA	15.5	6.69	02/25/2019	15:26
7	K1605-01DUP	1	Water	NA	NA	15.6	6.70	02/25/2019	15:27
8	CCV2	1	Water	NA	NA	22.2	12.02	02/25/2019	15:29

LB 101003

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph k1605

WorkList ID : 122691

Date : 02-25-2019 15:02:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/25/2019	Water	K1605-01-17	pH	Cool 4 deg C	PSEG01	K31	N48966	02/25/2019	9040C

Date/Time 02-05-19 15:10
Received by: Jo
Relinquished by: CL

Date/Time 02-15-19 16:00
Received by: Jo
Relinquished by: CL