



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB101010

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	02/25/2019	17:35
2	CAL2	1	Water	NA	NA	21.3	7.00	02/25/2019	17:36
3	CAL3	1	Water	NA	NA	21.3	10.02	02/25/2019	17:37
4	ICV	1	Water	NA	NA	21.2	6.99	02/25/2019	17:38
5	CCV1	1	Water	NA	NA	21.2	2.01	02/25/2019	17:40
6	K1348-08	1	Solid	20.02	20	19.3	8.14	02/25/2019	17:41
7	K1348-08DUP	1	Solid	20.03	20	19.4	8.15	02/25/2019	17:42
8	K1608-01	1	Solid	20.05	20	20.6	6.77	02/25/2019	17:43
9	K1610-02	1	Solid	20.02	20	21.5	7.62	02/25/2019	17:45
10	K1610-04	1	Solid	20.05	20	21.7	6.60	02/25/2019	17:46
11	K1610-06	1	Solid	20.03	20	21.4	7.51	02/25/2019	17:47
12	K1610-08	1	Solid	20.04	20	21.5	7.04	02/25/2019	17:48
13	K1610-10	1	Solid	20.02	20	20.9	8.10	02/25/2019	17:49
14	CCV2	1	Water	NA	NA	21.3	12.01	02/25/2019	17:50

18101010

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity K1610

WorkList ID : 122690

Date : 02-25-2019 15:02:16

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/25/2019	Solid	K1348-08	Corrosivity	Cool 4 deg C	PSEG05		OK-02-022519	02/25/2019	9045D
02/22/2019	Solid	K1608-01	Corrosivity	Cool 4 deg C	CLEA11	N11	OIL-WATER-SEPERATOR-1	02/22/2019	9045D
02/25/2019	Solid	K1610-02	Corrosivity	Cool 4 deg C	PSEG01	K31	WV-A	02/25/2019	9045D
02/25/2019	Solid	K1610-04	Corrosivity	Cool 4 deg C	PSEG01	K31	WV-B	02/25/2019	9045D
02/25/2019	Solid	K1610-06	Corrosivity	Cool 4 deg C	PSEG01	K31	WV-C	02/25/2019	9045D
02/25/2019	Solid	K1610-08	Corrosivity	Cool 4 deg C	PSEG01	K31	WV-D	02/25/2019	9045D
02/25/2019	Solid	K1610-10	Corrosivity	Cool 4 deg C	PSEG01	K31	WV-E	02/25/2019	9045D

Date/Time 02-25-19 15:20
 Received by: Jo
 Relinquished by: CL

Date/Time 02-25-19 18:00
 Received by: 30
 Relinquished by: CL