



PERCENT SOLID

Supervisor: apatel
Analyst: AMANDEEP
Date: 2/26/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:30
In Date: 02/25/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 108
Time OUT: 09:30
Out Date: 02/26/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %SOLIDS-OVEN

QC:LB101011

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K1348-07	OK-02-022519	1	1.14	9.98	9.3	92.3
K1594-21	B-1 (4-5)	2	1.15	9.76	8.1	80.7
K1594-22	B-1 (8-9)	3	1.13	9.59	8.05	81.8
K1594-24	B-2 (8-9)	4	1.15	9.96	8.81	86.9
K1594-26	B-2B (21-22)	5	1.12	9.96	8.56	84.2
K1594-28	B-3 (8-9)	6	1.11	9.97	8.7	85.7
K1594-29	B-4 (4-5)	7	1.14	9.93	9.02	89.6
K1594-31	B-4B (14-15)	8	1.15	9.9	8.47	83.7
K1594-34	B-6 (9-10)	9	1.15	9.76	8.66	87.2
K1594-38	B-8 (7.5-8.5)	10	1.12	9.92	8.95	89.0
K1594-39	B-9 (4-5)	11	1.13	9.9	9.39	94.2
K1594-40	B-9 (7.5-8.5)	12	1.16	9.81	8.23	81.7
K1610-01	WV-A	13	1.15	9.98	8.44	82.6
K1610-03	WV-B	14	1.14	9.96	8.54	83.9
K1610-05	WV-C	15	1.16	9.72	8.24	82.7
K1610-07	WV-D	16	1.17	9.73	8.65	87.4
K1610-09	WV-E	17	1.14	9.97	6.48	60.5
K1613-01	R-1	18	1.15	9.74	8.61	86.8
K1613-02	R-1A	19	1.14	9.78	8.69	87.4
K1613-03	L-1	20	1.18	9.93	8.84	87.5
K1613-04	L-1A	21	1.19	9.7	8.67	87.9
K1620-01	SB-01	22	1.14	9.76	9.12	92.6
K1620-02	SB-01	23	1.14	9.76	9.12	92.6
K1620-03	SB-02	24	1.15	9.86	8.16	80.5
K1620-04	SB-02	25	1.15	9.86	8.16	80.5
K1620-05	SB-03	26	1.15	9.79	7.47	73.1
K1620-06	SB-03	27	1.15	9.79	7.47	73.1
K1620-07	SB-04	28	1.18	9.96	8.87	87.6
K1620-08	SB-04	29	1.18	9.96	8.87	87.6
K1620-09	SB-05	30	1.13	9.83	8.78	87.9
K1620-10	SB-05	31	1.13	9.83	8.78	87.9

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$