



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB101023

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.6	02/26/2019	11:55
2	K1605-02		1	8	>100.0	>212.0	>212.0	02/26/2019	12:20
3	K1605-02DUP		1	8	>100.0	>212.0	>212.0	02/26/2019	12:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB101023

WorkList Name : FP22619

WorkList ID : 122734

Date : 02-26-2019 11:14:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/28/2019	Water	K1605-02	C Flash Point	Cool 4 deg C	PSEG01	K31	N48966	02/25/2019	1010A

Date/Time 02/26/19 11:20
 Received by: KS
 Relinquished by: CP

Date/Time 02/26/19 12:55
 Received by: CP
 Relinquished by: KS