



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB101026

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP71996
COD calibration std. 100 ppm	WP71994
COD calibration std. 10 ppm	WP71992
COD calibration std. 150 ppm	WP71995
COD calibration std. 50 ppm	WP71993
COD calibration std. 0 ppm	WP71990
COD CCV std, 50ppm	WP71997
COD LOD std, 5ppm	WP71998
COD LOQ std, 10.0PPM	WP71999
COD Digestion Vials Low Level 0-150Mg/L	W2468

Temp In (C): 150	Date In: 02/26/2019	Time In: 11:00
Temp Out (C): 150	Date Out: 02/26/2019	Time Out: 13:01

Intercept: 0.2078

Slope: 0.9999

Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	02/26/2019	14:00
2	CAL2	10	1	Water	11.000	02/26/2019	14:00
3	CAL3	50	1	Water	50.000	02/26/2019	14:01
4	CAL4	100	1	Water	99.000	02/26/2019	14:01
5	CAL5	150	1	Water	151.000	02/26/2019	14:02



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB101026

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	49.797	02/26/2019	14:02
2	ICB				1	Water	0.000	-0.208	02/26/2019	14:03
3	CCV1	50			1	Water	50.000	49.797	02/26/2019	14:03
4	CCB1				1	Water	0.000	-0.208	02/26/2019	14:04
5	LB101026BL				1	Water	0.000	-0.208	02/26/2019	14:04
6	LB101026BS	50			1	Water	50.000	49.797	02/26/2019	14:05
7	K1559-06				10	Water	40.000	39.796	02/26/2019	14:05
8	K1560-07				1	Water	7.000	6.793	02/26/2019	14:06
9	K1560-08				1	Water	13.000	12.793	02/26/2019	14:06
10	K1565-12				10	Water	35.000	34.796	02/26/2019	14:07
11	K1605-01				1	Water	59.000	58.798	02/26/2019	14:07
12	K1605-01DUP				1	Water	60.000	59.798	02/26/2019	14:08
13	K1605-01MS	50			1	Water	110.000	109.803	02/26/2019	14:08
14	K1605-01MSD	50			1	Water	111.000	110.803	02/26/2019	14:09
15	CCV2	50			1	Water	51.000	50.797	02/26/2019	14:09
16	CCB2				1	Water	0.000	-0.208	02/26/2019	14:10